



July 30th, 2018

RFQ for procurement of an Infrared-Variable Angle Spectroscopic Ellipsometer

This is an RFQ (Request for Quote) for procurement of Infrared-Variable Angle Spectroscopic Ellipsometer as part of a limited tender for the Centre for Nano Science and Engineering (CeNSE) at IISc, Bangalore.

CeNSE is a multidisciplinary research department at IISc that houses a 14,000 sq. ft. cleanroom and characterization facility used by 50 faculty members from various disciplines at IISc. CeNSE also runs a program called Indian Nanoelectronics Users Program (INUP) which has allowed 4200 participants from more than 700 universities and institutes all over India to use the facilities at CeNSE. Consequently, any tool in CeNSE receives significant exposure to scientific community at IISc and beyond. The vendors are requested to factor in the value of this exposure in to their quotes. Details of existing facilities and INUP program can be gleaned from:

<http://nnfc.cense.iisc.ac.in/>

<http://www.mncf.cense.iisc.ac.in/>

<https://www.inup.cense.iisc.ac.in/>

Also, CeNSE hosts equipment on behalf of vendors, as a national standard or 'model' system. If the vendor is interested, CeNSE can consider working out an arrangement for the Spectroscopic Ellipsometer.

Procedure

1. Vendors will be required to submit a technical proposal and a commercial proposal in **two separate sealed envelopes**. Only vendors who meet the technical requirement will be considered for the commercial negotiation.
2. **The deadline for submission of proposals is the 15th of August 2018, 5:30 pm Indian Standard Time.** Proposals should arrive at the Main office, GF-15, Centre for Nano Science and Engineering, Indian Institute of Science, Bangalore 560012, India, by the above deadline.
3. The decision of purchase committee will be final.
4. The technical proposal should contain a compliance table with 5 columns. The first column must list the technical requirements, in the order that they are given in the technical configuration below. The second column should describe your compliance in a "Yes" or "No" response. If "No" the third column should provide the extent of the deviation (please provide quantitative responses). The fourth column should state the reasons for the deviation, if any. The fourth column can be used to compare your tool with that of your competitors or provide details as requested in the technical requirements table below.
5. Any additional capabilities or technical details, that you would like to bring to the attention of the purchase committee, can be listed at the end of the technical table.
6. Vendors are encouraged to highlight the advantages of their tools over comparable tools from the competitors

7. If multiple systems can fulfill the requirements, vendors can submit multiple bids.
8. In the commercial bid, please provide itemized cost of the system and *required* accessories, such as software, power supply, etc.
9. As an option, please provide itemized cost for any *suggested* accessories/add-ons that may enhance the usability, capability, accuracy or reliability of the tool. Vendors are encouraged to quote for as many add-ons as their tool portfolio permits.
10. The quotes should be CIF Bangalore, India. So please include cost of shipping, insurance and other cost.
11. Please indicate the warranty provided with the tool. Warrant of 3 years of more is preferred.
12. Provide itemized cost for *required* spares for 2 years of operation. For sake of this calculation, the vendor may assume active tool usage of 20 hours/ week. This number will be used to estimate the life cycle cost of the tool.
13. Clarify if periodic (preventive) maintenance be done by a trained on-site engineer or requires a specialist from the OEM.
14. If maintenance requires OEM, as an additional option, provide cost of an annual maintenance contract (AMC) for 3 years, post warranty. The AMC must cover 1 scheduled and 1 emergency visit per year. It must also indicate who will service the AMC, an Indian agent or the OEM. The AMC cost must also include an itemized list of spares that are essential for the scheduled visits.
15. The RFQ must include references of 3 previous installations, preferable in India. Please provide the names and contact addresses of the referees, so that the committee can contact them independently.
16. The vendor should have installed at least 2 ellipsometer system in India. . Please provide the names and contact addresses of the referees, so that the committee can contact them independently.
17. Any questions can be directed to Prof. Shankar Kumar Selvaraja, TF-04, Centre for Nano Science and Engineering, Indian Institute of Science, Bangalore 560012, India. (shankarks@iisc.ac.in)

Technical Requirements

1.	Primary Requirement	<ul style="list-style-type: none">• Ability to perform Variable Angle Infra-red Spectroscopic Ellipsometry without modular changes• Spectral range: 1 micron to 25 micron
2.	Process capability	<ul style="list-style-type: none">• Automatic incidence angle control: 12.5 – 90 deg C• Micro-spot optics
3.	Tools specifications	<ul style="list-style-type: none">• Polarizer/compensator and Analyser arms, Rotating compensatory, Automated Attenuators, Light source module, independent adjustment of analyser and polarizer arms• Spectral resolution: Variable from 0.4 to 32 cm⁻¹• Spectral Accuracy: $\pm 0.2\text{cm}^{-1}$• Lamp and detector lifetime details should be provided

4.	Standard mode	<ul style="list-style-type: none"> At the time of installation, all relevant standard material model files should be provided
5.	Sample stage	<ul style="list-style-type: none"> Standard sample dimensions upto 6 inch Standard manual sample stage Fully automatic motorized sample stage for full wafer mapping as a option
6.	Detector specifications	<ul style="list-style-type: none"> DTGS atmospheric detector with the range into Infrared
7.	Calibration sample	<ul style="list-style-type: none"> Calibration sample to be provided along with the sample data
8.	Footprint & weight	<ul style="list-style-type: none"> The system should compatible with better than class 1000 cleanroom environment. Please specify the total foot print in cm x cm, and weight.
9.	Process software	<ul style="list-style-type: none"> Software for stage control and integrated data analysis Three copies of the software and soft and hard copy manual Software upgrade support period Lifetime model support from the vendor
10	Periodic Maintenance	<ul style="list-style-type: none"> The system should require minimal maintenance. Mention the recommended preventive maintenance schedule for the system. Any accessories needed for periodic preventive maintenance for 3 years e.g. lamps, should be mentioned in separately the itemized quote. Can the preventive maintenance be done by a trained on-site engineer or requires a specialist from the OEM? If the latter, please provide cost of a 3 year AMC consumables. The system should be supported by a trained local representative and should have a 48hour window of response
11	Installation and Training	<ul style="list-style-type: none"> Installation and training at customer site, by the experts from principals should be part of the package. During the installation all the specifications of the measurements should be verified for acceptance by the customer. If periodic maintenance can be done by the on-site engineer, please include the cost of training the engineer.
12	Power & utilities	<ul style="list-style-type: none"> The instrument should work with Indian standards Mention the power requirement. Mention any utility requirement (water, air, exhaust, etc.)
13	Safety	<ul style="list-style-type: none"> Mention any special safety requirement of the tool

		<ul style="list-style-type: none">• The tool must come with a complement of interlocks to prevent common user errors.
14	Recommendation	<ul style="list-style-type: none">• The system must submit references from atleast 3 previous installations• The names and contact addresses of the referees must be submitted with the proposal, so the purchase committee can contact them independently.
15	Pre-purchase testing	<ul style="list-style-type: none">• To ensure the equipment conforms for specifications, the committee requires the vendor to perform some standard tests <i>before</i> the purchase process is complete. The validity of the tender will hinge on the successful and accurate measurement of these test samples.• The vendor must conclude the testing and submit the data within 1 week of receipt of samples.
16	Acceptance tests	<ul style="list-style-type: none">• System will be accepted after confirming the thickness measurement from the installed system with cross-section or step height measurement.

Thanking you,

Shankar Kumar Selvaraja, Ph.D.
Assistant Professor
Centre for Nano Science and Engineering
Indian Institute of Science, Bangalore, India 560012.
Office : +91-80-2293-3342
E-mail: shankarks@iisc.ac.in