

Request for quote (RFQ) from domestic (India-based) manufacturers,
Indian OEM or its authorized Indian distributor

Summary

1.	Tender Number	OC/UM/2021/UV-Vis_01
2.	Tender Date	15th April 2021
3.	Item Description	Double beam UV-VIS Spectrophotometer
4.	Tender Type	Two bid system: (a) Technical Bid (Part A) (b) Commercial Bid (Part B)
5.	Place of tender submission	Department Office, room 44 Department of Organic Chemistry Indian Institute of Sciences, Bengaluru 560012
6.	Last Date & Time for submission of tender	28 th April 2021, 5:00 PM

To whom it may concern

This is a Request for quote (RFQ) from domestic (India-based) manufacturers, Indian OEM or its authorized Indian distributor only for procurement of **Double beam UV-VIS Spectrophotometer** at the department of Organic Chemistry (OC), Indian Institute of Science, Bangalore.

All interested vendors shall submit a response demonstrating their capabilities to produce the requested equipment to the primary point of contact listed below.

With respect to this tender, the rules laid out by the Government of India in order No. P45021/2/2017-pp-BE-II issued by the Public Procurement Section, Department or Promotion of Industry and Internal Trade, Ministry of Commerce and Industry, dated 4th June 2020 will be followed. As per this order, the government has defined a 'Class-I local supplier' as "a supplier or service provider whose goods, services or work offered for procurement, has local content equal to or more than 50%". A 'Class-II local supplier' is "a supplier or service provider, whose goods, services or works offered for procurement, has local content more than 20% but less than 50%". Only Class-I and Class-II local suppliers are eligible to participate in this open domestic tender. **Any "Non-local supplier" i.e. "a supplier or service provider, whose goods, services or works offered for procurement, has local content less than 20%" is ineligible to participate in this tender.**

The deadline for submission of proposals is 28th April 2021 by 5:00 PM. Proposals should arrive at the Department office, room 44, Department of Organic Chemistry, Indian Institute of Science, Bangalore, Karnataka 560012, India.

Direct all questions concerning the acquisition to Prof. Uday Maitra by email only at: maitra@iisc.ac.in

General Terms and Conditions

1. Quote should come only from Indian Original Equipment Manufacturer (OEM) or their authorized Indian distributor.
2. The bid should be submitted in the two-cover system, i.e. technical bid and commercial bid separately in sealed covers. The technical bid should contain all commercial terms and conditions, except the price.
3. The technical bid must contain a point-by-point technical compliance document. The technical proposal should contain a compliance table that should describe your compliance in a "yes" or "no" response against each of the items in the table listed in this RFQ. If the response is "no", the second column should state the extent of deviation. The third column should state the reason for the deviation, if any. The fourth column can be used to compare your tool with that of your competitors or provide details as requested in the technical requirement table below.
4. In the commercial bid, the price (in INR) should be inclusive of all discounts.
5. Provide certificates for country origin of manufacturing for each line items. Price of every line item in the commercial bid should be quoted along with the total quoted price for the instrument to be operational (installed and ready to use) in our facility.

6. The vendor should have qualified technical service personnel for the equipment based in India (preferably in Bangalore).
7. The covering letter should clearly state that whether the vendor is a Class-I or Class-II local supplier. Failing this the bid will be automatically rejected.
8. The vendor to state the percentage of the local content and provide self-certification that the item offered meets the minimum local content requirement. They should also give details of the location(s) at which the local value addition is made.
9. The lead time for the delivery of the equipment should not be more than 3 months from the date of receipt of our purchase order. It should be clearly mentioned in the technical and commercial bids.
10. All the quotations must be valid for at least 90 days at the time of submission.
11. List of customers and references: The Bidder should have supplied similar equipment in Central Universities preferably in centrally Funded Technical Institutes (IITs, IISC, IISER, NIT). Please provide the details and contact information.
12. The Bidder must not be blacklisted/banned/suspended or have a record of any service-related dispute with any organization in India or elsewhere. A declaration to this effect should be provided.
13. Items in addition to that listed in the technical table that you would like to bring to our attention, such as data sheets, technical plots etc. can be listed at the end of the compliance table.
14. Vendors are encouraged to highlight the advantage of their tools over comparable tools from the competitors.
15. If needed, a meeting for any technical clarifications can be scheduled with the undersigned by sending an email.
16. The Institute reserves the right to accept or reject any bid, or to annul the bidding process and reject all bids, at any time prior to the award of contract without thereby incurring any liability of the affected bidder or bidders.
17. Warranty terms and additional warranty options is a must for all the components. Please specify the service plan like whether the local distributor will address the issue or the parent company.
18. Terms and conditions for the annual maintenance contract beyond the warranty period should be mentioned.
19. After the award of purchase order, the vendor must provide an Order Acknowledgement within 30 days from the receipt of the Purchase Order.
- 21 Please quote the price of each optional line item separately. **The quotations should be on FOR-IISc Bangalore basis in INR only.**

Technical requirements: Please note that the requirements listed below are only guidelines. Vendors are requested to quote for equipment that meet the criteria to the best extent possible and list deviations, if any. Deviations are NOT an automatic reason for disqualification. They will be discussed by a technical group prior to making an informed decision.

Double Beam UV-Visible Spectrophotometer – technical specifications

Items	Specifications
Optics	Dual beam monochromator
Wavelength Range	185 to 800/900 nm with option of increase to 1400 nm with integrating sphere
Spectral Bandwidth	0.1-5 nm
Stray light	Less than 0.001% (220 nm NaI, 340 nm NaNO ₂)
Wavelength accuracy	±0.1 nm
Wavelength Repeatability	±0.1 nm
Scan Speed	60-6000 nm/min
Photometric Range	-6 to +6 Abs.
Photometric Accuracy	Less than 0.005 Abs (1.0 Abs)
Light Source(s)	Tungsten-Halogen and Deuterium Lamps
Noise level	Less than 0.00005 (500 nm)
Baseline Stability	Less than 0.0005 Abs/hr
Lamp interchange wavelength increment	Less than 0.1 nm
Detector	Industry standard detector (specify if there are options)
Thermoelectric temperature control (appropriate cooling system/circulator to be included)	0 – 110°C with stirrer
Operating temperature/humidity	10-35 °C/10-80% relative humidity (non-condensing)
Branded Fast Computer with 24" monitor	Specify the brand(s) for the PC/Monitor. The PC should be able to fully control the system.
Film holder attachment	Specify sample size minimum/maximum
Integrating sphere	Angle of incidence of light: 0 to 8 degree Wavelength range 220- 1400 nm
Powder sample holder for integrating sphere	Standard (specify details)
Micro sample holder for integrating sphere (1+1)	Sample size: 5-10 mm square, 0.5-2 mm thick
Software	<ul style="list-style-type: none"> • Real-time display of measurement with parameters • Adjust graph scale or use auto-scale • Automatically analyze data after measurements (specify capabilities) • Overlay spectral waveforms • Conversion (smoothing, differentiation, etc.) • Loadable on another PC for data processing

	<ul style="list-style-type: none"> Color analysis, color measurement software will also be required
10 mm path length quartz cells with stopper	Two matched pairs, one pair about 3 mL and the second pair < 1 mL capacity
Connectors/essential accessories	All cables and connectors to be provided for full operation of the system.
Multi cell holder (give as option)	
Additional warranty, AMC (give as option)	2 years and 2+2 years extended warranty (list separately)

Other requirements:

1.	Compatible operating system(s) for the interface software should be specified. Suitable software drivers available should be specified.
2.	Please include other options currently available which can be added on in the future.
3.	Installation and Training: Should be provided free of cost.
4.	The cost of shipping to IISc should be included.
5.	List of acceptance tests for on-site (vendor) inspection and after installation at IISc.
6.	A set of basic experiments for performing routine checks of acceptable operation with clear instructions to be provided. A standard sample for checking the absorbance should be provided.
7.	The payment terms will be specified in the commercial proposal which should be consistent with IISc's domestic purchase policies.
8.	Please provide details of the number of trained personnel in India, number in southern region or in Bangalore who can service the instrument.

Prof. Uday Maitra
 Department of Organic Chemistry
 Indian Institute of Science
 Bangalore, Karnataka 560012

maitra@iisc.ac.in