

## Corrigendum and updates related to the following tender invitation:

[https://iisc.ac.in/wp-content/uploads/2022/07/Global-Tender\\_AFM\\_MRC\\_2022\\_v1.pdf](https://iisc.ac.in/wp-content/uploads/2022/07/Global-Tender_AFM_MRC_2022_v1.pdf)

Please refer to Notice Inviting Global Tender dated July 25, 2022 for Supply and installation of atomic force microscope with various scanning probe capabilities at MRC, IISc Bangalore.

Further to the same following revisions are provided which need to be considered while submission of tender bids:

Listed Tender Specification / Clause	Revised Tender Specification / Clause
Section 4 – Clause 13.7 – Sub Clause 1 xxii) Quantitative Nano Scale Mapping more(QNM)	Section 4 – Clause 13.7 – Sub Clause 1 xxii) Quantitative Nano Scale Mapping more(QNM) is preferred or under Optional accessories in the listed specifications and may be offered as an optional item accordingly.
Section 4 – Clause 13.7 7. Controller should have 3 or more in-built locking amplifier.	Section 4 – Clause 13.7 7. Controller should have 2 or more in-built locking amplifier.
Section 4 – Clause 13.8 2. The system must include an active vibration isolation table. Not preferred integrated table or enclosure, must supply separate vibration table	Section 4 – Clause 13.8 2. The system must include a suitable vibration isolation table which is compatible for use with the offered system and demonstration of listed capabilities and noise specifications. It is preferable to have an integrated table & enclosure.
Section 4 – Clause 12.10 (to be read as 13.10) 1. Must include two-year warranty on all parts and labor.	Section 4 – Clause 13.10 1. Must include two-year warranty on all parts and labor. Additional extended warranty for year 3 should be mentioned separately as optional for each extended year.
Section 4 – Clause 12.10 ( to be read as 13.10) 4. At least minimum 10 User AFM software should be provided with instrument.	Section 4 – Clause 13.10 4. At least minimum 6 User AFM software licenses should be provided with instrument. Additional licenses per user should be offered as an option.

The above changes / clarifications are considered as the part and parcel of the tender document and shall be binding on all the parties.

Last date of submission: 15 August, 2022