

# Global Tender

This is a Request for Quote (RFQ) from the Indian Institute of Science (IISc), Bangalore, for the supply and installation of “Liquid Chromatograph - Quadrupole Time-of-Flight High Resolution Mass Spectrometer (HRMS) with Ion Mobility” at the Divisional Mass Spectrometry Facility, Biological Sciences Building, IISc, Bangalore.

This instrument is exempted from local tendering as per the reference (List. No. 105), Mass Spectrometers (including combinations with Liquid /Gas Chromatography) in the GOI office memorandum dated 28<sup>th</sup> June 2024, No. F. 4/1/2023 – PPD (Pt.).

1	Section 1	Bid Schedule	
2	Section 2	Eligibility Criteria	As specified by IISc
3	Section 3	Terms and Conditions	As specified by IISc
4	Section 4	Specifications	Technical specifications
5	Section 5	Technical Bid	Annexure 1: Bidder details
			Annexure 2: Declaration regarding experience of bidder
			Annexure 3: Declaration regarding clean track record of the bidder
			Annexure 4: Declaration of acceptance of tender
			Annexure 5: Terms and conditions. Details of item quoted
6	Section 6	Commercial bid	Quotation with Price, Technical specifications of the Equipment

Statutory condition: The committee can cancel the tender at any time.

## Notice Inviting Tender

### SECTION 1 – BID SCHEDULE

1	Tender No	IISc/DBS/Mass Spectrometry-02/2024
2	Tender Date	05.09.2024
3	Item Description	Supply and Installation for “Liquid Chromatograph - Quadrupole Time-of-Flight High Resolution Mass Spectrometer (HRMS) with Ion Mobility” at the Divisional Mass Spectrometry Facility, Biological Sciences Building, IISc, Bangalore.
4	Tender Type	Two Bid System a) Technical Bid (Part-A) b) Commercial Bid (Part-B)
5	Place of Submission and tender opening	Prof. Jayanta Chatterjee, Convenor, Divisional Mass Spectrometry Facility, Biological Sciences Building, Indian Institute of Science, Bangalore – 560012, India
6	Last Date & Time for submission of tender	30.09.2024 before 17: 00 hrs (21days)
7	For further clarifications	Prof. Jayanta Chatterjee, Convenor, Divisional Mass Spectrometry Facility, Biological Sciences Building, Indian Institute of Science Bangalore- 560012, India Email: <a href="mailto:jayanta@iisc.ac.in">jayanta@iisc.ac.in</a> with a copy to: <a href="mailto:mharshitha@iisc.ac.in">mharshitha@iisc.ac.in</a>

## **SECTION 2 – ELIGIBILITY CRITERIA**

### **Prequalification criteria:**

- 1. The Bidder's firm should have existence for a minimum of 3 years. (Enclose Company Registration Certificate)**
- 2. The Bidder should have qualified technical service personnel for the instrument(s) based in India.**
- 3. If the Bidder is a local distributor/dealer/Agent, it is mandatory to attach authorization certificate along with the bid from the original equipment manufacturer.**
- 4. The bidder should sign and submit the declaration for Acceptance of Terms and Conditions as per - Annexure 4**
- 5. The Bidder must not be blacklisted/banned/suspended or have a record of any service-related dispute with any organization in India or elsewhere. A declaration to this effect has to be given as per Annexure 3.**

## SECTION 3 – TERMS AND CONDITIONS

### A) Submission of Tender:

1. All documentations in the tender should be in English.
2. Tender should be submitted in two envelopes (two bid system).
  - a. Technical Bid (Part-A) – Technical bid consisting of all technical details and check list for conformance to technical specifications.

The technical proposal should contain a technical compliance table with 5 columns.

- i. The first column must list the technical requirements, in the order that they are given in the technical requirement below.
  - ii. The second column should provide specifications of the instrument against the requirement. Please provide quantitative responses wherever possible.
  - iii. The third column should describe your compliance with a “Yes” or “No” only. Ensure that the entries in column 2 and column 3 are consistent.
  - iv. The fourth column should state the reasons/explanations/context for deviations, if any.
  - v. The fifth column can contain additional remarks from the OEM. You can use this opportunity to highlight technical features, qualify response of previous columns, or provide additional details, compare your solution with that of your competitors or provide details as requested in the technical requirements table below.
- b. Commercial Bid (Part-B) –  
Indicating item wise price for the items mentioned in the technical bid, as per the format of quotation provided in tender, and other commercial terms and conditions
3. The technical bid and price bid should each be placed in a sealed cover, superscripting on both the envelopes the tender no. and the due date and both these sealed covers are to be placed in a bigger cover which should also be sealed and duly superscripted with the Tender No & Due Date.
  4. The SEALED COVER superscripting tender number / due date & should reach to **Prof. Jayanta Chatterjee, Convenor, Divisional Mass Spectrometry Facility, Biological Sciences Building, Indian Institute of Science, Bangalore – 560012**, India on or before due date mentioned in the tender notice. In case due date happens to be holiday the tender will be accepted and opened on the next working day. If the quotation cover is not sealed, it will be rejected.
  5. All queries are to be addressed to the person identified in “Section 1 – Bid Schedule” of the tender notice.
  6. The items are required for research purposes, and IISc is a DSIR registered institution, hence eligible for GST exemption (i.e. GST @ 5%). While submitting the price quote, this point must be taken care. For getting GST exemption certificate, successful bidders must submit, a formal request together with Invoice copy and Purchase order copy.
  7. GST/other taxes, levies etc., are to be indicated separately. The BIDDER should mention GST Registration and PAN in the tender document (Indian Bidders only).
  8. If price is not quoted in Commercial Bid as provided in tender document the bid is liable to be rejected.
  9. The Institute reserves the right to accept or reject any bid and to annul the bidding process and reject all bids at any time prior to the award of contract, without there by incurring any liability to the affected bidder or bidders or any obligation to inform the affected bidder or bidders.
  10. Incomplete bids will be summarily rejected.

- B) Cancellation of Tender:  
Notwithstanding anything specified in this tender document, IISc Bangalore, in its sole discretion, unconditionally and without having to assign any reason, reserves the rights:
- a. To accept OR reject lowest tender or any other tender or all the tenders.
  - b. To accept any tender in full or in part.
  - c. To reject the tender, offer not confirming to the tender terms.
- C) Validity of the Offer:  
The offer shall be valid 90 Days from the date of opening of the commercial bid.
- D) Evaluation of Offer:
1. The technical bid (Part A) will be opened first and evaluated.
  2. Bidders meeting the required eligibility criteria as stated in Section 2 of this document shall only be considered for Commercial Bid (Part B) opening. Further, agencies not furnishing the documentary evidence as required will not be considered.
  3. Pre-qualification of the bidders shall not imply final acceptance of the Commercial Bid. The agency may be rejected at any point during technical evaluation or during commercial evaluation. The decision in regard to acceptance and / or rejection of any offer in part or full shall be the sole discretion of IISc Bangalore, and decision in this regard shall be binding on the bidders.
  4. The award of contract will be subject to acceptance of the terms and conditions stated in this tender.
  5. Any offer which deviates from the vital conditions (as illustrated below) of the tender is liable to be rejected:
    - a. Non-submission of complete offers.
    - b. Receipt of bids after due date and time and or by email / fax (unless specified otherwise).
    - c. Receipt of bids in open conditions.
  6. In case any BIDDER is silent on any clauses mentioned in these tender documents, IISc Bangalore shall construe that the BIDDER had accepted the clauses as of the tender and no further claim will be entertained.
  7. No revision in the terms and conditions quoted in the offer will be entertained after the last date and time fixed for receipt of tenders.
  8. Lowest bid will be calculated based on the total price of all items tendered for Basic equipment along with accessories for installation, operation, pre-processing and post processing, optional items, recommended spares, warranty, annual maintenance contract. Also see section G.
- E) Pre-requisites:  
The bidder will provide the prerequisite installation requirement of the equipment along with the technical bid.
- F) Warranty:  
Two years on-site comprehensive warranty (year wise breakup value should be shown in the commercial bid) including free supply of consumables, for all parts of the MS and accessories supplied and data analysis software from the date of functional installation. Free maintenance and service must be provided during warranty. Must include minimum 3 preventive maintenance visits per year. Complete warranty for the High Resolution QTOF with Ion Mobility MS to be quoted which includes both spares and labour.  
If the instrument is found to be defective, it has to be replaced or rectified at the cost of the bidder within 30 days from the date of receipt of written communications from IISc, Bangalore. If there is any delay in replacement or rectification, the warranty period should be correspondingly extended.
- G) Annual Maintenance Contract:  
An additional one-year annual maintenance (AMC) support should be provided on completion of warranty period. If the equipment cost is A, AMC is B, the lowest bid will be calculated as  $L1=A+5*B$ .

H) Purchase Order:

- A The order will be placed on the bidder whose bid is accepted by IISc based on the terms & conditions mentioned in the tender document.
- B The quantity of the items in tender is only indicative. IISc, Bangalore reserves the right to increase /decrease the quantity of the items depending on the requirement.
- C If the quality of the product and service provided is not found satisfactory, IISc, Bangalore reserves the right to cancel or amend the contract.

I) Delivery, Installation and Training:

The bidder shall provide the lead time to delivery, installation and made functional at IISc, Bangalore from the date of receipt of purchase order. The system should be delivered, installed and made functional within 90 days from the date of receipt of purchase order. The supply of the items will be considered as effected only on satisfactory installation and inspection of the system and inspection of all the items and features/capabilities tested by the IISc, Bangalore. After successful installation and inspection, the date of taking over of entire system by the IISc, Bangalore shall be taken as the start of the warranty period. No partial shipment is allowed.

The bidder should also arrange for technical training to the local facility technologists and users.

J) Payment Terms:

The payments to non-domestic vendors will be through a Letter of Credit and milestone of the payment will be determined after the mutual discussions with the successful bidder. As per GFR no advance payment can be made to domestic vendors, unless an equal amount of bank guarantee is provided.

K) Statutory Variation:

Any statutory increase in the taxes and duties subsequent to bidder's offer, if it takes place within the original contractual delivery date, will be borne by IISc, Bangalore subject to the claim being supported by documentary evidence. However, if any decrease takes place the advantage will have to be passed on to IISc, Bangalore.

L) Disputes and Jurisdiction:

Any legal disputes arising out of any breach of contract pertaining to this tender shall be settled in the court of competent jurisdiction located within the city of Bangalore, India.

M) General:

- A All amendments, time extension, clarifications etc., within the period of submission of the tender will be communicated electronically. No extension in the bid due date/time shall be considered on account of delay in receipt of any document(s) by mail.
- B The bidder may furnish any additional information, which is necessary to establish capabilities to successfully complete the envisaged work. It is however, advised not to furnish superfluous information.
- C The bidder may visit the installation site before submission of tender, with prior intimation.
- D All imported equipment should be quoted in the currency of the country of origin, and all locally sourced items should be quoted in Indian Rupees.
- E Any information furnished by the bidder found to be incorrect, either immediately or at a later date, would render the bidder liable to be debarred from tendering/taking up of work in IISc, Bangalore.

**Section – 4 – Technical Specifications**

**A) Technical Specification for Liquid Chromatograph - Quadrupole Time-of-Flight High Resolution Mass Spectrometer (HRMS) with Ion Mobility**

Sl.no	Specifications	Essential/Desirable	Remarks (By Vendor)
1	<p><b>Proposed applications and features:</b></p> <ul style="list-style-type: none"> <li>• The system is required for qualitative and quantitative analysis, such as discovery proteomics, complex biofluids such as plasma and serum analysis, quantitative proteomics, and targeted and non-targeted metabolomics workflows. It should include software to compare and analyze metabolite data from the mass spec of (metabolomics) blood, serum, plasma, urine, and routine samples. The instrument should be the latest and top-most model.</li> <li>• The system should be upgradeable with software for all usage, including available databases for analyzing proteomics and metabolomics workflow studies.</li> <li>• The instrument must be available with Ion Mobility separation for separating isobaric species, and every ion generated in the system must pass through a gas and provide a Collisional-Cross-Section (CCS) of every ion. The ion mobility must offer an additional degree of separation to provide greater selectivity and greater spectral clarity of all components of a sample injection.</li> </ul>	Essential	
2	<p><b>Specifications for ESI QTOF HRMS with Ion Mobility Spectrometry</b></p>		
	<p><b>System Hardware Specifications</b>            A High-Resolution Mass Spectrometer comprising a Quadrupole for isolation of various mass-to-charge ratios, Time of flight tube, and Ion mobility for obtaining accurate mass information up to 4 decimal places</p>		
3	<p><b>Mass analyzer</b></p>		
	<ul style="list-style-type: none"> <li>• The HRMS instrument must be a high resolution QTOF with high sensitivity and high scan speed</li> <li>• The instrument must contain an Ion mobility spectrometry analyser delivering high mobility resolution</li> <li>• The system should be capable of performing intact protein mass identification, peptide mass fingerprinting, and precursor ion selection followed by high energy fragmentation to generate high resolution MS/MS spectra</li> <li>• High sensitivity (pmol to amol range) fragmentation at high sensibility and High energy (CID)</li> <li>• The mass analyzer should have a quadrupole followed by a Collision cell &amp; a high resolution TOF analyzer for both MS and MS/MS analysis</li> <li>• The QTOF HRMS with Ion Mobility should have following minimum specifications:               <ul style="list-style-type: none"> <li>• Quadrupole Mass Range: 20 to 20,000 m/z or better</li> <li>• Acquisition Rate: 30 Spectra per second or better in MS &amp; MS/MS mode</li> <li>• Resolution: The resolution of the instrument should be 50,000 (FWHM) at a scan speed of 30 spectra/sec in both MS and MS/MS mode or better</li> <li>• Mass Accuracy: 1 ppm for both MS &amp; MS/MS modes</li> </ul> </li> </ul>		

	<ul style="list-style-type: none"> <li>• The analyser / detections system must have the capability of producing data with a dynamic range of up to 4 orders of magnitude</li> <li>• Ion Mobility aligned separation and fragmentation for understanding the post-translational modifications like phosphorylation, acetylation, disulfide scrambling etc for protein sample for proteomics</li> </ul>		
<b>4</b>	<b>ESI Ionization Source</b>		
	<ul style="list-style-type: none"> <li>• The instrument should have provision for two different ESI sources – 1) Nano Flow &amp; 2) Analytical flow applications covering proteomics and metabolomics workflows.</li> <li>• Positive and negative ionization capabilities must be included as standard on the instrument.</li> <li>• An isolation valve must be fitted to the source to allow the source elements to be removed and cleaned without breaking instrument vacuum, maximizing instrument uptime.</li> <li>• The ionization sources should be Plug and play probes.</li> </ul>		
<b>5</b>	<b>Ion Mobility</b>		
	<ul style="list-style-type: none"> <li>• A Third dimension of molecular separation based on ion mobility for the separation of isobaric species, isolation of challenging co-eluting contaminants and reduction of high background noise, improving data quality and enhancing selectivity for challenging samples (cell lysates ; serum ; plasma ; biopsies ; tissues) requiring advanced analytical separations should be available within the instrument</li> <li>• The user should be able to switch ON &amp; OFF this mode through the software itself to switch between routine Qtof mode &amp; IMS-Qtof mode</li> <li>• The system should have ability to select m/z before IMS</li> <li>• IMS should provide a unique method to enhance the characterization of molecular structure through the determination of collision cross section values (CCS)</li> </ul>		
<b>6</b>	<b>Modes of operation</b>		
	<ul style="list-style-type: none"> <li>• MS Scanning</li> <li>• MS/MS product Ion Scanning</li> <li>• Simultaneous MS &amp; MS/MS scanning</li> <li>• TOF- MRM/PRM mode for performing quantitation at sub ppb levels</li> <li>• DDA/IDA scanning. The acquisition of data should also be possible by setting thresholds for ion intensity for both MS &amp; MSMS and by switching between the two modes</li> <li>• The acquisition software should allow for automated MS/MS, IMS/MS, IMS/MS/MS, DIA, IMS/MS/MS (CID/IMS, IMS/CID and CID/IMS/CID) and data directed acquisitions</li> <li>• The operation should have internal mass correction technique for better mass accuracy</li> <li>• The instrument should be capable of internal reference mass correction for MS and MS/MS operation without losing sensitivity</li> <li>• Direct Infusion: Syringe pump or equivalent for direct infusion of samples</li> </ul>		
<b>7</b>	<b>Data station with software</b>		
	<ul style="list-style-type: none"> <li>• A workstation should be provided for controlling the mass spectrometer, for data acquisition &amp; data processing and analysis</li> <li>• Automated mass calibration, auto tuning, resolution and sensitivity check should be performed by software</li> <li>• The data processing software must incorporate an elemental composition calculator as standard</li> </ul>		



	<ul style="list-style-type: none"> <li>• Software tools to address screening, component identification and structure elucidation tools</li> <li>• The software should be capable of providing true isotropic patterns and fragment ions</li> <li>• Deconvolution software for the determination of molecular weights of high molecular weight compounds</li> </ul>		
<b>8</b>	<b>Ultra-High Performance Liquid Chromatograph (UHPLC/UPLC)</b>		
	<ul style="list-style-type: none"> <li>• Binary pump designed for UHPLC/UPLC with vacuum degassing capability of solvents</li> <li>• Flow rate: 0.010 - 2.00 ml/min or better</li> <li>• Mobile phase Accessories, inline filters, plus dampers, mixers and filters</li> <li>• Column Temperature Control should be 20 °C to 80 °C</li> <li>• Pressure should be 15,000 psi or better</li> <li>• Autosampler should be capable of holding 90 vials or more</li> <li>• Sample Temperature: 4-40 °C, programmable in 0.1 °C increments</li> <li>• The vendor should provide the 500 number of sample vials</li> <li>• The vendor should provide 2 columns suitable for proteomics and metabolomics workflow</li> </ul>		
<b>9</b>	<b>Warranty and Training</b>		
	<ul style="list-style-type: none"> <li>• Two years on-site comprehensive warranty for all parts of the MS and accessories supplied. Free maintenance and service must be provided during warranty. Must include minimum 3 preventive maintenance visits per year. An additional one-year annual maintenance (AMC) support should be provided. Complete warranty for the High Resolution QTOF with Ion Mobility MS to be quoted which includes both spares and labor</li> <li>• A hands-on training should be provided by the vendor application team on the instrument operation, basic instrument maintenance and core applications level training</li> <li>• In case of breakdown or any malfunction of the instrument, service should be provided by the vendor within 48 hours of the reporting time</li> </ul>		
<b>10</b>	<b>Standards and other requirements</b>		
	<ul style="list-style-type: none"> <li>• Necessary standards and reagents for installation should be provided</li> <li>• The vendor should issue an undertaking that the availability of spare parts for at least 10 years from the date of successful installation</li> <li>• The quoted system should be compatible with free software upgrade for at least 10 years from the date of successful installation</li> <li>• Users lists of similar equipment supplied in India should be provided</li> </ul>		

**B) Training and demonstration:**

**At least 2 or 3 training programs on usage of the machine (hardware and software) must be demonstrated by the successful bidder at bidder's cost to the end users at IISc, Bangalore**

## **Section 5: Technical Bid**

**The technical bid should furnish all requirements of the tender along with all annexure in this section and submitted to**

**Prof. Jayanta Chatterjee,  
Convenor,  
Divisional Mass Spectrometry Facility,  
Biological Sciences Building,  
Indian Institute of Science  
Bangalore- 560012, India**

**Annexure-1:**

**Details of the Bidder**

The bidder must provide the following mandatory information & attach supporting documents wherever mentioned:

<b>Details of the Bidder</b>		
<b>SI No</b>	<b>Name of the Bidder</b>	
<b>1</b>	<b>Nature of Bidder (Attach attested copy of Certificate of Incorporation/ Partnership Deed)</b>	
<b>2</b>	<b>Registration No/ Trade License, (attach attested copy)</b>	
<b>3</b>	<b>Registered Office Address</b>	
<b>4</b>	<b>Address for communication</b>	
<b>5</b>	<b>Contact person- Name and Designation</b>	
<b>6</b>	<b>Telephone No</b>	
<b>7</b>	<b>Email ID</b>	
	<b>Website</b>	
<b>8</b>	<b>PAN No. (attach copy)</b>	
<b>9</b>	<b>GST No. (attach copy)</b>	

**Signature of the Bidder**

**Name**  
**Designation, Seal**

**Date:**

**Annexure-2:**

**Declaration regarding experience**

**To,**

**Prof. Jayanta Chatterjee,  
Convenor,  
Divisional Mass Spectrometry Facility,  
Biological Sciences Building,  
Indian Institute of Science  
Bangalore- 560012, India**

**Ref: Tender No: XXXXXXXXX**

**Dated: XXXXX**

**Supply and Installation of "Liquid Chromatograph - Quadrupole Time-of-Flight High Resolution Mass Spectrometer (HRMS) with Ion Mobility" at Divisional Mass Spectrometry Facility, Biological Sciences Building.**

**Sir,**

**I've carefully gone through the Terms & Conditions contained in the above referred tender. I hereby declare that my company / firm has ---- years of experience in supplying and Installation of Multicolour Benchtop Flow Cytometer/Analyzer.**

**(Signature of the Bidder)**

**Printed Name**

**Designation, Seal**

**Date:**

**Annexure-3:**

**Declaration regarding track record**

To,

**Prof. Jayanta Chatterjee,  
Convenor,  
Divisional Mass Spectrometry Facility,  
Biological Sciences Building,  
Indian Institute of Science  
Bangalore- 560012, India**

**Ref: Tender No: XXXXXXXX**

**Dated: XXXXX**

**Supply and installation "Liquid Chromatograph - Quadrupole Time-of-Flight High Resolution Mass Spectrometer (HRMS) with Ion Mobility" at Divisional Mass Spectrometry Facility, Biological Sciences Building.**

Sir,

I've carefully gone through the Terms & Conditions contained in the above referred tender. I hereby declare that my company/ firm is not currently debarred / blacklisted by any Government / Semi Government organizations / institutions in India or abroad. I further certify that I'm competent officer in my company / firm to make this declaration.

Or

I declare the following

Sl.No.	Country in which the company is Debarred /blacklisted / case is Pending	Blacklisted / debarred by Government / Semi Government/Organizations /Institutions	Reason	Since when and for how long
--------	-------------------------------------------------------------------------	------------------------------------------------------------------------------------	--------	-----------------------------

(NOTE: In case the company / firm was blacklisted previously, please provide the details regarding period for which the company / firm was blacklisted and the reason/s for the same).

Yours faithfully

(Signature of the Bidder)

Name

Designation, Seal

Date:

**Annexure – 4:**

**Declaration for acceptance of terms and conditions**

**To,**

**Prof. Jayanta Chatterjee,  
Convenor,  
Divisional Mass Spectrometry Facility,  
Biological Sciences Building,  
Indian Institute of Science  
Bangalore- 560012, India**

**Ref: Tender No: XXXXXX**

**Dated: XXXX**

**Supply and installation of “Liquid Chromatograph - Quadrupole Time-of-Flight High Resolution Mass Spectrometer (HRMS) with Ion Mobility” at Divisional Mass Spectrometry Facility, Biological Sciences Building.**

**Sir,**

**I've carefully gone through the Terms & Conditions as mentioned in the above-referred tender document. I declare that all the provisions of this tender document are acceptable to my company. I further certify that I'm an authorized signatory of my company and am, therefore, competent to make this declaration.**

**Yours faithfully,**

**(Signature of the Bidder)**

**Name**

**Designation, Seal**

**Date:**

**Annexure – 5:**

**Details of items quoted:**

<b>Company Name</b>	
<b>Product Name</b>	
<b>Part / Catalogue number</b>	
<b>Product description / main features</b>	
<b>Detailed technical specifications</b>	
<b>Remarks</b>	

**Instructions to bidders:**

- 1. Bidder should provide technical specifications of the quoted product/s in detail.**
- 2. Bidder should attach product brochures along with technical bid.**
- 3. Bidders should clearly indicate compliance or non-compliance of the technical specifications provided in the tender document.**

### Section 6: COMMERCIAL BID

The commercial bid should be furnished with all requirements of the tender with supporting documents as mentioned under:

S.No	Description	Cat. Number	Quantity	Unit Price	Sub total
1.	Essential items noted in the technical specification				
1.a	... (details of essential items)				
1.b	...				
2.	Optional items noted in the technical specification				
2.a	... (details of essential items)				
2.b	...				
3.	Accessories for operation and installation				
4.	All Consumables, spares and software to be supplied locally				
5.	Warranty (2 years) Two years on-site comprehensive warranty (year wise breakup value should be shown in the commercial bid) including free supply of consumables, for all parts of the MS and accessories supplied and data analysis software from the date of functional installation. Free maintenance and service must be provided during warranty. Must include minimum 3 preventive maintenance visits per year. Complete warranty for the High Resolution QTOF with Ion Mobility MS to be quoted which includes both spares and labour.				
6.	AMC An additional one-year annual maintenance (AMC) support should be provided on completion of warranty period. If the equipment cost is A, AMC is B, the lowest bid will be calculated as $L1=A+5*B$ .				
7.	Cost of Insurance and Airfreight				
8.	CIP/CIF IISc, Bengaluru				

Any additional items

S.No	Description	Cat. Number	Quantity	Unit Price	Sub total

**Addressed to:**

**Prof. Jayanta Chatterjee, Convenor,  
Divisional Mass Spectrometry Facility,  
Biological Sciences Building, Indian Institute of Science  
Bangalore- 560012, India**



## **SECTION 7 – CHECK LIST**

**(This should be enclosed with technical bid- Part A)**

**The following items must be checked before the Bid is submitted:**

**1. Sealed Envelope “A”: Technical Bid**

**1. Section 5- Technical Bid (each page signed by the authorized signatory and sealed) with the below annexures:**

- a. Annexure 1: Bidders details**
- b. Annexure 2: Declaration regarding experience**
- c. Annexure 3: Declaration regarding clean track record**
- d. Annexure 4: Declaration for acceptance of terms and conditions**
- e. Annexure 5: Details of items quoted**

**2. Copy of this tender document duly signed by the authorized signatory on every page and sealed.**

**2. Sealed Envelope “B”: Commercial Bid**

**Section 6: Commercial Bid**

**Your quotation must be submitted in two envelopes: Technical Bid (Envelope A) and Commercial Bid (Envelope B) super scribing on both the envelopes with Tender No. and due date and both of these in sealed covers and put in a bigger cover which should also be sealed and duly super scribed with Tender No., Tender description & Due Date.**