

DEPARTMENT OF MECHANICAL ENGINEERING
INDIAN INSTITUTE OF SCIENCE (IISc), BENGALURU, INDIA

Local Tender Notice

Tender Notification Ref No.: ME/ENQ/TNDR/KV/AK/25-26/01

Date: 01.01.2025

The Department of Mechanical Engineering, Indian Institute of Science Bengaluru, invites tenders for supply of “IR Camera”. This Invitation for Bids is open to 2all domestic (India based) manufacturers, Indian OEM or its authorised Indian distributors only.

The scope of the supply includes Installation, Commissioning and Training at site for the “IR Camera”.

Tender Summary

1	Tender number	<i>ME/ENQ/TNDR/KV/AK/25-26/01</i>
2	Tender Date	<i>01.01.2025</i>
3	Item Description	<i>IR Camera</i>
4	Tender Type	Two Bid System: (a) Technical Bid (Part A) (b) Commercial Bid (Part B)
5	Place of Tender submission	Prof. Alope Kumar Associate Professor Prof. Koushik Viswanathan Associate Professor, Department of Mechanical Engineering, Indian Institute of Science, Bangalore - 560012
6	Last date & Time for submission of tender	<i>22nd Jan 2025 @ 5.00 P.M</i>

To whom it may concern

This is a **Request for quote (RFQ)** from **Indian Agencies** for supply including Installation, Commissioning, training at site for “IR Camera” at the “**Department of Mechanical Engineering (ME), Indian Institute of Science, Bengaluru.**”

This Invitation for Bids is open to only domestic (India based) manufacturers, Indian OEM or its authorized Indian distributors. All interested vendors shall submit a response demonstrating their capabilities to produce the requested equipment to the primary point of contact listed below. ***With respect to this tender, the rules laid out by the Government of India in order No. P45021/2/2017-pp-BE-II issued by the Public Procurement Section, Department or Promotion of Industry and Internal Trade, Ministry of Commerce and Industry, dated 4th June 2020 will be followed.***

The deadline for submission of proposals is **22nd Jan 2025 @ 5.00 P.M** Proposals should arrive at the office of

**The Chair,
Department of Mechanical
Engineering,
Indian Institute of Science,
Bengaluru - 560012,
Karnataka, India.**

*Direct all questions concerning the acquisition to addresses to . Prof. Alope Kumar at alokekumar@iisc.ac.in Associate Professor and **Prof. Koushik Viswanathan** at: koushik@iisc.ac.in*

General Terms and Conditions

1. The bid should be submitted in the two-cover system, i.e., technical bid and commercial bid separately in sealed covers. The technical bid should contain all commercial terms and conditions, except the price.
2. The technical bid must contain a point-by-point technical compliance document. The technical proposal should contain a compliance table that should describe your compliance in a "yes" or "no" response against each of the items in the table listed in this RFQ. If "no" the second column should state, the extent of deviation. The third column should state the reason for the deviation, if any. The fourth column can be used to compare your tool with that of your competitors or provide details as requested in the technical requirement table below.
3. In the commercial bid, the price should be inclusive of all discounts.
4. The quotations should be on FOR-IISc Bangalore basis in INR only.
5. The vendor should have qualified technical service personnel for the equipment based in India (preferably in Bangalore).The Bidder should belong to either Class-1 or Class-2 suppliers distinguished by their “local content” as defined by recent edits to GFR. They should mention clearly which class they belong to in the cover letter. a) Class-1 supplier: Goods and services should have local content of equal to or more than 50%. b) Class-2 supplier: Goods and services should have local content of equal to or more than 20% and less than 50%.
6. The covering letter should clearly state that whether the vendor is a Class-I or Class-II local supplier. Failing this the bid will be automatically rejected.
7. Any “Non-local supplier” i.e. “a supplier or service provider, whose goods, services or works offered for procurement, has local content less than 20%” is ineligible to participate in this tender.
8. Bidders offering imported products will fall under the category of non-local suppliers. They cannot claim themselves as Class-1 local suppliers/Class-2 local suppliers by claiming the services such as transportation, insurance, installation, commissioning, training, and other sales service support like AMC/CMC, etc., as local value addition.
9. Purchase preference as defined by the recent edits to GFR (within the “margin of purchase preference”) will be given to the Class-1 supplier.

10. MSMEs can seek an exemption to some qualification criteria. IISc follows GFR2017 for such details.
11. The vendor to state the percentage of the local content and provide self-certification that the item offered meets the minimum local content requirement. They should also give details of the location(s) at which the local value addition is made.
12. The lead time for the delivery of the equipment should not be more than 3 months from the date of receipt of our purchase order. It should be clearly mentioned in the technical and commercial bids.
13. All the quotations must be valid for at least 90 days at the time of submission.
14. List of customers and references: **The Bidder should have supplied similar equipment in Central Universities, preferably in centrally Funded Technical Institutes (IITs, IISC, IISER, NIT). Please provide the details and contact information.**
15. The Bidder must not be blacklisted/banned/suspended or have a record of any service-related dispute with any organization in India or elsewhere. A declaration to this effect should be provided.
16. Items in addition to that listed in the technical table that you would like to bring to the attention of the committee, such as data sheets, technical plots etc. can be listed at the end of the compliance table.
17. Vendors are encouraged to highlight the advantage of their Laser optics and Delivery Head over comparable Laser optics and Delivery Head from the competitors.
18. If needed, a meeting for any technical clarifications can be scheduled with the undersigned by sending an email.
19. The Institute reserves the right to accept or reject any bid, or to annul the bidding process and reject all bids, at any time prior to the award of contract without thereby incurring any liability of the affected bidder or bidders.
20. Warranty terms and additional warranty options is a must for all the components. Please specify the service plan like whether the local distributor will address the issue or the parent company.
21. Terms and conditions for the annual maintenance contract beyond the warranty period should be mentioned.
22. After the award of purchase order, the vendor must provide an Order Acknowledgement within 30 days from the receipt of the Purchase Order.
23. Please quote the price of each optional line item, separately.

Desired Technical requirements

Please note that the requirements listed below are only guidelines. It does not disbar bids that do not meet the criteria listed. Vendors are requested to quote for equipment that meet the criteria to the best extent possible and list deviations. Deviations are NOT an automatic reason for disqualification. They will be discussed by the technical committee prior to making an informed decision.

Technical Specifications: IR Camera

A. Camera:

1. Detector Type: (InSb) to allow high sensitivity
2. Spectral Range: 3.0 – 5.0 μm for high dynamic range
3. Resolution: 640 x 512
4. Detector Pitch: 25 μm
5. Thermal Sensitivity/NETD: ≤ 20 mK
6. Well Capacity: 11.0 M electrons
7. Operability: $\geq 99.95\%$
8. Sensor Cooling: Closed cycle rotary
9. Readout Type: Snapshot
10. Readout Modes: Asynchronous integrate while read Asynchronous integrate then read
11. Synchronization Modes: Sync-in, Sync-out
12. Integration Time: 270 ns to 687 sec
13. Frame Rate (Full Window) Programmable; 0.0015 Hz to 520 Hz
14. Sub window Mode: Flexible windowing down to 32 x 4
15. Dynamic Range: 14-bit
16. On-Camera Image Storage: RAM (volatile): atleast 16 GB, up to 26000 frames, full frame SSD (non-volatile): ≥ 4 TB capability, with atleast 512 GB supplied
17. Radiometric Data Streaming: Simultaneous Gigabit Ethernet (GigE Vision), Camera Link
18. Standard Video: HDMI
19. Command and Control: GigE, Camera Link
20. Filter wheel with 4 filters
21. Lens: f/2.5, 50mm band matched to detector
22. Temperature calibration: 10°C to 350°C, with accuracy $\pm 2\%$. Higher temperature calibration (upto 2000 C) to be included in quote
23. Ambient drift compensation with factory calibration, atleast 4 temperatures should be monitored to implement this and displayed in software
24. Camera should be IP67 rated
25. Calibration should be done without using any moving parts like apertures
26. User calibration should be allowed
27. Super framing with presets of integration time to cover wide temperature range in one exposure without needing any moving parts like apertures
28. Software should be usable with both cooled and uncooled cameras
29. Software should have ability to stream data and perform real-time analysis
30. Software should support MP4 video export and text format (CSV) frame export
31. SDK support to be provided for custom solution development
32. Quality system standard: ISO 9001:2015 or better
33. Warranty of 24 months that covers camera and all provided accessories must be provided

B. Accessories:

1. Lens (50 mm):
 - a. Band matched to detector
 - b. Aperture f/2.5 or higher
 - c. Transmission $>90\%$
 - d. Minimum focus distance not more than 50cm
 - e. Calibrated from 0 to 350C
2. Heavy duty tripod (Al or SS) with 2-axis fluid head for mounting
3. 1 Transport case for storing and transporting camera

Additional details:

C. Prequalification of Bidder

1. The Vendor must be a reputed original equipment manufacturer (OEM) of the infrared Thermography system having support network in India OR must be an authorized agent / supplier of the Original Equipment Manufacturer. The vendor must also be registered with GOVT. OF India and should have very good support network in India.
2. The Vendor should submit evidence of satisfactory performance in India of at least 05 (five) cooled infrared Thermography systems manufactured by the same OEM, during the last 2 years.
3. OEM (Original Equipment Manufacturer) must have office in India with calibration checking facility. Camera must be repairable in India.
4. The detector has to be supplied within 3 months from date of PO, failing which the order will be cancelled.

- ✚ Should have proven record of successful installations within Indian education/research institutes. Proof of such installation must be enclosed.
- ✚ The OEM/Supplier should have trained service engineers stationed in India for any on-site service requirement, details to be provided in the offer.
- ✚ OEM/supplier should have requisite stock of necessary spare parts in India.

TERMS AND CONDITIONS

1. Warranty period: 1 year.

2. Supplier Credibility:

- a. The Bidder/Vendor Must have supplied minimum 3 similar orders to reputed Government Organizations in INDIA in the past 3 years. Copies of Purchase Orders to be enclosed along with the Technical Bid as proof.
- b. Supplier should compulsorily indicate details of facilities / expertise/ qualification of support staff in India. Factory trained engineer/s should be available in India for complete product support
- c. Please enclose User list in INDIA.
- d. Minimum 3 Reference letters of similar system supplied in INDIA need to be submitted for further consideration.

3. Publications:

- a. As our Research Work is of critical nature, Vendors need to enclose reference publications/application note on the usage of “ IR Camera” to show expertise of the product being offered.
- b. As a scope of future work, we intend to use planar measurements on the same set-up. Vendors can provide reference publications of using “ IR Camera” (from same OEM).

4. Institute reserves the right for final selection of items.

5. Vendors may quote for any other items/accessories separately as “Optional Items”.

6. Company should have their own production facility in India and a strong sales/service record in Bangalore.

For queries or clarifications, please contact:

Prof. Alope Kumar and Prof. Koushik Viswanathan
Department of Mechanical Engineering
Indian Institute of Science,
Bengaluru, Karnataka 560012
alokekumar@iisc.ac.in
koushik@iisc.ac.in

Annexure-I

Note: Compliance Certificate must be enclosed with the Technical bid. Non submission of Compliance Certificate will lead to disqualification of the bidder.

Technical Specifications: IR Camera

A. Camera:

1. Detector Type: (InSb) to allow high sensitivity
2. Spectral Range: 3.0 – 5.0 μm for high dynamic range
3. Resolution: 640 x 512
4. Detector Pitch: 25 μm
5. Thermal Sensitivity/NETD: ≤ 20 mK
6. Well Capacity: 11.0 M electrons
7. Operability: $\geq 99.95\%$
8. Sensor Cooling: Closed cycle rotary
9. Readout Type: Snapshot
10. Readout Modes: Asynchronous integrate while read Asynchronous integrate then read
11. Synchronization Modes: Sync-in, Sync-out
12. Integration Time: 270 ns to 687 sec
13. Frame Rate (Full Window) Programmable; 0.0015 Hz to 520 Hz
14. Sub window Mode: Flexible windowing down to 32 x 4
15. Dynamic Range: 14-bit
16. On-Camera Image Storage: RAM (volatile): atleast 16 GB, up to 26000 frames, full frame SSD (non-volatile): ≥ 4 TB capability, with atleast 512 GB supplied
17. Radiometric Data Streaming: Simultaneous Gigabit Ethernet (GigE Vision), Camera Link
18. Standard Video: HDMI
19. Command and Control: GigE, Camera Link
20. Filter wheel with 4 filters
21. Lens: f/2.5, 50mm band matched to detector
22. Temperature calibration: 10°C to 350°C, with accuracy $\pm 2\%$. Higher temperature calibration (upto 2000 C) to be included in quote
23. Ambient drift compensation with factory calibration, atleast 4 temperatures should be monitored to implement this and displayed in software
24. Camera should be IP67 rated
25. Calibration should be done without using any moving parts like apertures
26. User calibration should be allowed
27. Super framing with presets of integration time to cover wide temperature range in one exposure without needing any moving parts like apertures
28. Software should be usable with both cooled and uncooled cameras
29. Software should have ability to stream data and perform real-time analysis
30. Software should support MP4 video export and text format (CSV) frame export
31. SDK support to be provided for custom solution development
32. Quality system standard: ISO 9001:2015 or better
33. Warranty of 24 months that covers camera and all provided accessories must be provided

B. Accessories:

1. Lens (50 mm):
 - a. Band matched to detector
 - b. Aperture f/2.5 or higher
 - c. Transmission $>90\%$
 - d. Minimum focus distance not more than 50cm
 - e. Calibrated from 0 to 350C
2. Heavy duty tripod (Al or SS) with 2-axis fluid head for mounting
3. 1 Transport case for storing and transporting camera

C-Comply, NC- Non Comply, D- Deviation

Annexure-II

MANUFACTURER'S AUTHORISATION FORM

[The bidder shall require the manufacturer to fill in this form in accordance with the instructions indicated. This letter of authorization should be on the letterhead of the Manufacturer and should be signed by the person with the proper authority to sign documents that are binding on the Manufacturer.]

Date: [insert date (as day, month and year) of Bid Submission]

Tender No.: [insert number from Invitation for Bids]

To: The Chair, Department of Mechanical Engineering, IISc, Bengaluru 560012

We [*insert complete name of Manufacturer*], who are official manufacturers of [*insert full address of Manufacturer's factories*], do hereby authorize [*insert complete name of Bidder*] to submit a bid the purpose of which is to provide the following Goods, manufactured by us [*insert name and or brief description of the Goods*], and to subsequently negotiate and sign the Contract.

We hereby extend our full guarantee and warranty with respect to the Goods offered by the above firm.

Signed: [insert signature(s) of authorized representative(s) of the Manufacturer]

Name: [insert complete name(s) of authorized representative(s) of the Manufacturer]

Title: [insert title]

Duly authorized to sign this authorization on behalf of: [insert complete name of Bidder]