

**Tender notification for Water Purification System for
Laboratory Water**

Tender Number: (MBU/RN/IISc/08-02/2025) Global Tender

1	Section 1	Bid Schedule	
2	Section 2	Eligibility Criteria	As specified by IISc
3	Section 3	Terms and Conditions	As specified by IISc
4	Section 4	Specifications	Technical Specifications
5	Section 5	Technical Bid	Annexure 1: Bidder details
			Annexure 2: Declaration regarding the experience of the bidder
			Annexure 3: Declaration of track record
			Annexure 4: Declaration of acceptance of terms and conditions
			Annexure 5: Details of items quoted.
6	Section 6	Commercial Bid	Quotation with price, and technical specifications of the equipment.

Section 1: Bid Schedule

1	Tender Number	MBU/RN/IISC/08-02/2025
2	Tender Date	14/08/2025
3	Item Description	Water Purification system for laboratory water
4	Tender Type	Two bid system: (a) Technical Bid (Part A) (b) Commercial Bid (part B)
5	Place of tender submission	Ms. Pavithra, MBU-Office Molecular Biophysics Unit Indian Institute of Science Bengaluru – 560012 email: chair.mbu@iisc.ac.in office.mbu@iisc.ac.in ; attn: Siddhartha P. Sarma
6	Last Date and Time for tender submission	15 th September 2025
7	Primary Point of Contact	Prof. Rishikesh Narayanan, Chairman, Molecular Biophysics Unit, Indian Institute of Science, Bangalore, 560012, Karnataka, India. Phone. +91-80-22932460/3552 Attn: Siddhartha P. Sarma (sidd@iisc.ac.in)

The procurement of **Water Purification System for Laboratory Water** at the office of Molecular Biophysics Unit (MBU) in the Indian Institute of Science, Bangalore. All interested vendors shall submit a response demonstrating their capabilities to produce the requested equipment to the primary point of contact.

Section 2: Eligibility Criteria

Prequalification criteria:

1. The bidder should sign and submit the declaration for Acceptance of Terms and Conditions as per -Annexure 4.
2. The Bidder must not be blacklisted/banned/suspended or have a record of any service- related dispute with any organization in India or elsewhere. A declaration to this effect has to be given as per Annexure 3.
3. System Catalogue should be produced with the Technical Bid. Original Invoice, Original Warranty Certificate, and Original Test Reports should be produced for all imported items from OEM (Original Equipment Manufacturer) at the time of supply of equipment.
4. Manufacturers should have ISO or equivalent international standard certificate. Please attach the required certificate with the bid.
5. The supplier will support the user with all the spares for a minimum period of 5 years.
6. Details of experienced service engineers including contact details should be provided in the tender document.

Section 3: Terms and conditions

(A) Submission of Tender

1. All documentation in the tender should be in English.
2. Vendors will be required to submit a technical proposal and a commercial proposal in **two separate sealed envelopes** (two bid systems). The technical bid should contain all commercial terms and conditions, except the price. **Only vendors who meet the technical requirement will be considered for the commercial discussion.**
3. **The technical bid** (Part A) must contain a point-by-point technical compliance document. The technical proposal should contain a compliance table with 5 columns.
 - a. The first column must list the technical requirements, in the order that they are given in the technical requirements below.
 - b. The second column must provide the specification of the instrument against the requirement (please provide quantitative responses wherever possible)
 - c. The third column should describe the compliance with a “YES” or “NO” only. Ensure that the entries in column 2 and column 3 are consistent.
 - d. The fourth column should clearly state the reasons/explanations/context for deviations if any.
 - e. The fifth column may contain additional remarks. It can be used to highlight the technical features, qualify the response of previous columns, or provide additional details.
4. **The commercial bid** (Part B) – Indicating item-wise price for the items mentioned in the technical bid, **as per the format of quotation provided in the tender**, and other commercial terms and conditions. The commercial bid should indicate the following separately: (a) equipment price (b) optional items (c) Shipping cost and (d) the Total cost.
5. The technical bid and price bid should each be placed in separate sealed covers, super scribing on both the envelopes a tender no. and the due date. Both these sealed covers are to be placed in a bigger cover which should also be sealed and duly superscripted with the Tender No, Tender Description & Due Date.
6. The SEALED COVER super scribing tender number / due date & should reach Department of Developmental Biology & Genetics, Indian Institute of Science, Bangalore – 560012, India on or before the due date mentioned in the tender notice. In case the due date happens to be a holiday the tender will be accepted and opened on the next working day. If the quotation cover is not sealed, it will be rejected.
7. All queries are to be addressed to the person identified in “Section 1: Bid Schedule” of the tender notice.
8. GST/other taxes, levies etc., are to be indicated separately. The BIDDER should mention

GST Registration and PAN in the tender document (Indian Bidders only).

9. In addition to that listed in the technical table that the vendor would like to bring to the attention of the committee, such as data sheets, technical plots etc. must be listed at the end of the compliance table. The committee will go through the data provided and those available in their website to evaluate the suitability.
10. The decision of the purchase committee will be final. The Institute reserves the right to accept or reject any bid or to annul the bidding process and reject all bids, at any time before the award of the contract without thereby incurring any liability of the affected bidder or bidders.
11. Only the Original Equipment Manufacturer or their authorized representatives across the globe shall participate in the bid.
12. The order will be placed only on the bidder who participated in the bid.

(B) Cancellation of Tender

Notwithstanding anything specified in this tender document, IISc Bangalore, in its sole discretion, unconditionally and without having to assign any reason, reserves the rights:

- a. To accept OR reject the lowest tender or any other tender or all the tenders.
- b. To accept any tender in full or in part.
- c. To reject the tender, offer not confirming to the tender terms.

(C) Validity of the Offer

The offer shall be valid for at least 90 Days from the date of opening of the commercial bid.

(D) Evaluation of the Offer

1. The technical bid (Part A) will be opened first and evaluated.
2. Bidders meeting the required eligibility criteria as stated in Section 2 of this document shall only be considered for Commercial Bid (Part B) opening. Further, agencies not furnishing the documentary evidence as required will not be considered.
3. Pre-qualification of the bidders shall not imply final acceptance of the Commercial Bid. The agency may be rejected at any point during technical evaluation or commercial evaluation. The decision regarding acceptance and/or rejection of any offer in part or full shall be the sole discretion of IISc Bangalore, and the decision in this regard shall be binding on the bidders.
4. The award of the contract will be subject to acceptance of the terms and conditions stated in this tender.
5. Any offer which deviates from the vital conditions (as illustrated below) of the tender is liable to be rejected:
 - a. Non-submission of complete offers.
 - b. Receipt of bids after the due date and time and or by email/fax (unless specified otherwise).
 - c. Receipt of bids in open conditions.
6. In case any BIDDER is silent on any clauses mentioned in these tender documents, IISc Bangalore shall consider that the BIDDER had accepted the clauses as of the tender and no further claim will be entertained.
7. No revision of the terms and conditions quoted in the offer will be entertained after the last date and time fixed for receipt of tenders.
8. The lowest bid will be calculated based on the total price of all items tendered for Basic equipment along with accessories selected for installation, operation, preprocessing and post-processing, optional items, recommended spares, warranty, and annual maintenance contract.

(E) Pre-requisites:

The bidder will provide the prerequisite installation requirement of the equipment along with the technical bid.

(F) Warranty:

The complete system is to be under a warranty period of **a minimum of 3 year** from the date of delivery, including a free supply of spare parts and online assistance. If the instrument is found to be defective, it must be replaced or rectified at the cost of the bidder within 30 days from the date of receipt of written communications from IISc Bangalore. If there is any delay in replacement or rectification, the warranty period should be correspondingly extended.

(G) Purchase Order:

1. The order will be placed on the bidder whose bid is accepted by IISc based on the terms & conditions mentioned in the tender document.
2. The quantity of the items in the tender is only indicative. IISc, Bangalore reserves the right to increase /decrease the quantity of the items depending on the requirement.
3. If the quality of the product and service provided is not found satisfactory, IISc Bangalore reserves the right to cancel or amend the contract.

(H) Delivery and Installation

The bidder shall provide the lead time to delivery, installation and made functional at IISc Bangalore from the date of receipt of the purchase order. The system should be delivered, installed, and made functional **within 3 months** from the date of receipt of the purchase order. The supply of the items will be considered as effected only on satisfactory installation and inspection of the system and inspection of all the items and features/capabilities tested by the IISc, Bangalore. After successful installation and inspection, the date of taking over the entire system by the IISc, Bangalore shall be taken as the start of the warranty period. No partial shipment is allowed. The technical installation personnel/engineer, who will perform the installation, must possess proof of a training certificate from the manufacturing company/factory/principal, which should be submitted along with the tender offer.

(I) Payment Terms:

100% payments (except AMC) will be released after completion of delivery and satisfactory installation subject to TDS as per rules. As per GFR, no advance payment can be made to the authorized Indian distributors unless an equal amount of bank guarantee is provided.

(J) Statutory Variation:

Any statutory increase in the taxes and duties after the bidder's offer, if it takes place within the

original contractual delivery date, will be borne by IISc, Bangalore subject to the claim being supported by documentary evidence. However, if any decrease takes place the advantage will have to be passed onto IISc, Bangalore.

(K) Disputes and Jurisdiction:

Any legal disputes arising out of any breach of contract about this tender shall be settled in the court of competent jurisdiction located within the city of Bangalore, India.

(L) General:

1. All amendments, time extensions, clarifications etc., within the period of submission of the tender will be communicated electronically. No extension in the bid due date/time shall be considered on account of delay in receipt of any document(s) by mail.
2. The bidder may furnish any additional information, which is necessary to establish capabilities to complete the envisaged work. It is, however, advised not to furnish superfluous information.
3. The bidder may visit the installation site before submission of the tender, with prior intimation.

Any information furnished by the bidder found to be incorrect, either immediately or at a later date, would render the bidder liable to be debarred from tendering/taking up of work in IISc Bangalore.

Section 4: Technical Specifications

Part A. Technical specifications for Water Purification System for Laboratory Water

Lab Water Purification System – Type I & II Water Quality

Standalone Type II System which can produce purified water with $>5 \text{ M}\Omega\cdot\text{cm}$ @25 °C resistivity with pre-treatment cartridge, Reverse Osmosis, followed by standalone Type-1 Water Purification System which can produce Ultrapure Water with $18.2 \text{ M}\Omega\cdot\text{cm}$ @25 °C resistivity with $\text{TOC} < 5 \text{ ppb}$.

Feed Water Specification:

Water Type: Portable water

Pressure: 2 bar min to 6 bar max (29 to 86 psi).

Conductivity: $10 \mu\text{S}$ to $2000 \mu\text{S}$

pH: 4 to 10

Hardness (as CaCO_3): $< 300 \text{ ppm}$

Silica concentration: $< 30 \text{ ppm}$

Carbon dioxide concentration (CO_2): $< 60 \text{ ppm}$

Langelier Saturation Index (LSI): < 0.3

Fouling Index (FI5) or Silt Density Index (SDI): 7

Stage:1: Prefiltration Unit

The Prefiltration unit must consist of: Booster pump [Centrifugal Type] with controller, Filter with housing, Chassis, Main socket and power cord, Pressure Gauge, Timer unit for cartridge replacement and other accessories. The prefiltration unit should also consist of 3- & 1-Micron depth filters to remove the particulate matter from the feed water. The unit should have the automatic low/high pressure cutoff mechanism. The Pre-filtration system should be manufactured by the same company which manufactures the Water Purification system and indigenous unit if quoted will not be considered. The complete product catalogue of the prefiltration unit from the manufacturer is mandatory to be uploaded, Individual filter quality certificate if only submitted will not be considered.

Stage: 2: Pure Water/Type-2 water Generation System

- Pre-treatment cartridge with Poly Phosphate [anti-scaling compound], 0.5-micron filter and activated carbon, with Radio Frequency Identification tag [RFID] for easy traceability [Bar code or any other technologies will not be considered] & to have water consumption data stored in the system memory. In addition, the activated carbon is impregnated with a small level of silver to prevent bacteria growth. Users should have the provision to go to the menu option and find out the main consumable's status and traceability.
- System should have Pump with unique temperature feedback mechanism so that it works not only in high feed water temperature situations but also in low and high feed water temperature situations. The benefit of the same needs to be explained/highlighted with the help of the flow schematic in the product broucher.
- Latest Polyamide Thin film composite membrane RO which should be capable of achieving 99% rejection of Ions, Organics, Particles. Automatic product flow regulation on feed water temperature between 10 and 35 °C (45 to 95 °F).
- RO Reject Water recovery: Automatic regulation between 45 and 75% based on water analysis parameters introduced in the controller (E.R.A technology).Also it should have conductivity cell before and after RO to measure performance of RO membrane.
- The System should have the Electro De-ionization module (EDI Module) & EDI module should

contain conductive carbon Beads in Cathode compartment to avoid the use of softener before the module. Online Resistivity cell should also be present after EDI module to check the final type-2 product water quality.

- Product water is sanitized as a final purification step by passing through a 254 nm UV lamp providing bacterial removal at 4 LRV (log reduction value).
- System should have 24/7 Real time remote monitoring compatible and should have the provision to connect with BMS. It should have touchscreen display with detail water quality monitoring of makeup, storage. System should be capable of archiving data through USB port up to 2 years. It should have complete graphical traceability and real time monitoring of consumables with attached RFID Tag.
- Information menu should contain all the details to understand the main system's performance, (Product water Resistivity and % of RO rejection, RO pressure, RO water quality, RO membrane efficiency % ion rejection, RO feed water conductivity and Permeate water conductivity etc.
- Storage Reservoir: Storage capacity should be 100 liters with material of construction of Polyethylene for storage Type II Water. It should be controlled by water purification unit. Tank should be fully drainable capable of collecting water from the bottom. The tank should have Opaque walls to block sunlight and to prevent algae development, should be Cylindrical in shape to minimize surface area in contact with water. The tank should have Conical bottom which will allow complete draining for cleaning and rinsing. The tank should also have front valve for manual dispense of pure water. The tank must be of Compact space-saving design. Vent filter should also be provided which is built with activated carbon to adsorb volatile organics, a soda-lime bed to remove CO₂, a Durapore® hydrophobic membrane for particle and bacteria retention.

Product Water Quality: Type-2 water

Should meet or exceed ASTM Type II water:

Resistivity: >5 MΩ.

TOC: < 30 ppb.

Flow Rate: 40 Liters/Hour.

Stage: 3: Ultra-Pure Water/Type-1 water Generation System

Type-1 System should have the following technology insight:

- Nanotechnology based Ion Exchange resin to achieve faster kinetics for ionic removals.
- The water system should have a built-in UV lamp with emission at 185 nm and 254 nm wavelength for bacterial germicidal effect & photooxidation of organics.
- Option for attaching various application specific packs simultaneously for multiple application like LC-MS, Bacteriology, GC-MS, Endocrine Disruptors, PCR, Cell Culture.
- System should have the provision to go for hibernation mode (when lab is closed for more than 3 days) to save the energy & reduce the risk of contamination when not in use.
- Touch Screen colour Display: Size -7" Touch Screen LCD Display.
- Display Home screen should contain the details of Water Quality, Dispensing Options, Tank Water Level, Menu, Recirculation, Options for Dispense Report and Alerts/Alarms if any.
- Display should have the Screen-saver options for saving the power.
- Display menu should have multiple menus for Consumables, Maintenance, Information, Settings.
- Consumables Menu should have the option for Real time Graphical Monitoring of consumable status. Presence of E-Sure tag in Type-1 system consumables for traceability and better inventory planning.

- Maintenance menu should contain self-guided wizard (Graphical representation) of Pak/Cartridges replacements steps which intern will help he users to perform the activity themselves without any intervention of Service engineers.
- Precise on-line conductivity monitoring having a 0.01 cm-1 cell constant and a 0.1° C sensitive thermistor. No of Conductivity/ Resistivity Cell should be 5 or above to ensure the Input output water quality at various filtration stages.
- In-line TOC measurement with 185nm lamp should have the options to show TOC values in display for every dispensing.
- System should have options for volumetric dispensing to manage different volume requirements from ml to Liters.
- System should have the options to manage at least 3 different flows (High, Mid, Low) based on the requirements.
- Dispensing gun should contain magnetic lock system with dispensing arm to avoid the risk of breakage due to handling error.
- It should have the provision for hands free dispensing to reduce the risk of contamination though hands or human intervention.
- System should have the options for connecting water dispenser (along with main units) in 4 different places based on the Lab layout. Also, there should be a provision to connect dispensing units complete remotely to save effective bench space & enhance ease of use.
- System should have the check and dispense lights options.
- As and when required entire dispensing history should be available that can provide information on water quality and system maintenance, which can be downloadable through a data cable or USB key.
- Service (Contracts) historical data should be managed through cloud base online solutions for easy access of data to the users in remote conditions.
- Service Support: Direct Service Support from the manufacturer only not by any third party.
- common email-id/toll free number to register service-related issues.
- Wide range of service offerings like on-site verification/calibration performed directly by manufacturers only.
- Certifications & Compliance report for Water System: CE, cUL, FCC, ISO® 9001 v. 2000- and ISO® 14001 or better, All certificates need to be submitted after installation of the equipment.

Product Water Quality: Type-1 water:

Resistivity: 18.2 MΩ·cm at 25°C,

TOC (ppb) ≤ 5 ppb,

Bacteria < 0.01 cfu/ml,

Flow rate: Up to 2L/min.

Water purification systems should be manufactured in an ISO® 9001 v. 2008 and ISO 14001-v. 2004- or better, registered manufacturing site. Should be certified for safety and electromagnetic compatibility (CE, cUL; FCC). Total Warranty of One Year from the date of installation. The product installation and demo should be given to all the users after installation without any additional cost. The installation, maintenance, service of the system should be carried out by the manufacturer owned service engineers and not by any third party, with relevant service engineer certification. Quotes are expected only from reputed companies who has excellent track record of selling Lab water purification systems, having their manufacturer's installation of Lab water purification systems of 100 or more systems in Karnataka over the last 3 years. The OEM manufacturer must have dedicated company owned service engineers based out of Bangalore to provide immediate service support and ensure that service support is extended as and when

required. There should be a dedicated toll-free e-mail ID/contact number to log in all the service queries. The Manufacturer must also have their central warehouse located in India to supply the spares/consumables on immediate basis.

Quotation should be enclosed with a compliance statement , List of users at least 20 nos, located in Bangalore and other sites in Karnataka. Available performance certificates may be enclosed.

For the Water Purification the participating firms must quote all-inclusive delivery prices and entire shipment must be insured from manufacturers' warehouse to IISC–MBU.

Important: Please note that the Water purification system should be match all the technical specifications in a detailed document in the technical bid.

The documents may be addressed to the Chair, Molecular Biophysics Unit, IISc (Attention: Prof. Siddhartha P Sarma), Indian Institute of Science, Bangalore 560 012. Last date for receiving queries: August 30th, 2025. Please email chair.mbu@iisc.ac.in and office.mbu@iisc.ac.in with CC to sidd@iisc.ac.in.

Section 5- Technical Bid

The technical bid should furnish all requirements of the tender along with all annexures in this section and submitted to

Prof. Rishikesh Narayanan
Chairman,
Molecular Biophysics Unit,
Indian Institute of Science,
Bangalore, 560012,
Karnataka, India.
Phone. +91-80-22932460/3552

Attn: Siddhartha P. Sarma (sidd@iisc.ac.in)

Annexure-1:

Details of the Bidder

The bidder must provide the following mandatory information & attach supporting documents wherever mentioned:

Details of the Bidder

Sl. No Items Details

1. Name of the Bidder
2. Nature of Bidder (Attach attested copy of Certificate of Incorporation/ Partnership Deed)
3. Registration No/ Trade License, (attach attested copy)
4. Registered Office Address
5. Address for communication
6. Contact person- Name and Designation
7. Telephone No
8. Email ID
9. Website
10. PAN No. (attach copy)
11. GST No. (attach copy)

Signature of the Bidder

Name, Designation, Seal

Date

Annexure-2:

Declaration regarding experience

To,
Prof. Rishikesh Narayanan
Chairman,
Molecular Biophysics Unit,
Indian Institute of Science,
Bangalore, 560012,
Karnataka, India.
Phone. +91-80-22932460/3552

Attn: Siddhartha P. Sarma (sidd@iisc.ac.in)

Ref: Tender No: **MBU/RN/IISc/08-02/2025**

Dated: 13/08/2025

Supply and installation of **Water Purification System for Laboratory Water** at Department of Molecular Biophysics Unit, IISc Bangalore.

Sir/Ma'am,

I've carefully gone through the Terms & Conditions contained in the above referred tender. I hereby declare that my company / firm has ---- years of experience in supplying and installing high speed centrifuges

(Signature of the Bidder)

Name, Designation, Seal

Date

Annexure-3:

To,
Prof. Rishikesh Narayanan
Chairman,
Molecular Biophysics Unit,
Indian Institute of Science,
Bangalore, 560012,
Karnataka, India.
Phone. +91-80-22932460/3552
Attn: Siddhartha P. Sarma (sidd@iisc.ac.in)

Declaration of the track record
Department of Molecular Biophysics unit,
Indian Institute of Science, Bengaluru 560012, India

Ref: Tender No: **MBU/RN/IISc/08-02/2025**

Dated: 14/08/2025

Supply and installation of **Water Purification System for Laboratory Water**, IISc
Bangalore.

Sir,
I've carefully gone through the Terms & Conditions contained in the above referred tender. I hereby declare that my company/ firm is not currently debarred /blacklisted by any Government / Semi Government organizations/institutions in India or abroad. I further certify that I'm competent officer in my company/firm to make this declaration.

Or

I declare the following,

Sl.No.	Country in which the company is Debarred/blacklisted	/case is Pending Blacklisted/debarred by Government/Semi Government/Organizations	/Institutions	Reasons	Since when and for how long
--------	--	---	---------------	---------	-----------------------------

(NOTE: In case the company / firm was blacklisted previously, please provide the details regarding the period for which the company / firm was blacklisted and the reason/s for the same).

Yours faithfully

(Signature of the Bidder)

Name Designation, Seal

Date

Annexure – 4:

Declaration for acceptance of terms and conditions

To,
Prof. Rishikesh Narayanan
Chairman,
Molecular Biophysics Unit,
Indian Institute of Science,
Bangalore, 560012,
Karnataka, India.
Phone. +91-80-22932460/3552
Email: office.mbu@iisc.ac.in
chair.mbu@iisc.ac.in

Attn: Siddhartha P. Sarma (sidd@iisc.ac.in)

Ref: Tender No: **MBU/RN/IISc/08-02/2025**

Dated: 14/08/2025

Subject- Supply and installation of **Water Purification System for Laboratory Water**, at

Molecular biophysics Unit", IISc Bangalore. Sir,
I've carefully gone through the Terms & Conditions as mentioned in the above referred tender document. I declare that all the provisions of this tender document are acceptable to my company. I further certify that I'm an authorized signatory of my company and am, therefore, competent to make this declaration.

Yours faithfully,

(Signature of the Bidder)

Name, Designation, Seal

Date

Annexure – 5:

Details of items quoted:

- a. Company Name
- b. Product Name
- c. Part / Catalogue number
- d. Product description / main features
- e. Detailed technical specifications
- f. Remarks

Instructions to bidders:

1. Bidder should provide technical specifications of the quoted product/s in detail.
2. Bidder should attach product brochures along with technical bid.
3. Bidders should clearly indicate compliance or non-compliance of the technical specifications provided in the tender document.

Section 6 – Commercial Bid

The bid should be furnished with all requirements of the tender with supporting documents as mentioned under:

Sl. No.	Description	Cat. Number	Quantity	Unit Price	Sub total
1.	Essential items noted in the technical specification				
1.a	... (details of essential items)				
1.b	...				
2.	Optional items noted in the technical specification				
2.a	... (details of optional items)				
2.b	...				
3.	Accessories for operation and installation				
4.	All Consumables, spares and software to be supplied locally				
5.	Warranty (1 year)				
6.	Cost of Insurance and Airfreight				

Any additional items

Sl. No	Description	Cat. Number	Quantity	Unit Price	Sub total

Addressed to

To,
Prof. Rishikesh Narayanan
Chairman,
Molecular Biophysics Unit,
Indian Institute of Science,
Bangalore, 560012,
Karnataka, India.
Phone. +91-80-22932460/3552
Email: office.mbu@iisc.ac.in
chair.mbu@iisc.ac.in

Attn: Siddhartha P. Sarma (sidd@iisc.ac.in)

Section 7 – Checklist

(This should be enclosed with technical bid- Part A)
The following items must be checked before the Bid is submitted:

1. Sealed Envelope “A”: Technical Bid

1. Section 5- Technical Bid (each page signed by the authorized signatory and sealed) with the below annexures:
 - a. Annexure 1: Bidders details
 - b. Annexure 2: Declaration regarding experience
 - c. Annexure 3: Declaration regarding clean track record
 - d. Annexure 4: Declaration for acceptance of terms and conditions
 - e. Annexure 5: Details of items quoted.
2. Copy of this tender document duly signed by the authorized signatory on every page and sealed.

2. Sealed Envelope “B”: Commercial Bid Section 6: Commercial Bid

Your quotation must be submitted in two envelopes: Technical Bid (Envelope A) and Commercial Bid (Envelope B) super scribing on both the envelopes with Tender No. and due date and both of these in sealed covers and put in a bigger cover which should also be sealed and duly super scribed with Tender No., Tender description & Due Date.

