



Date: 15.09.2025

Corrigendum

Tender Number: **CAF-IISc/IVC/2025-2026**

Tender Title: Supply and Installation of Individually Ventilated Cage (IVC) Units at the Central Animal Facility

Date of publishing: 03.09.2025

Corrigendum: 15.09.2025

Last date of submission: 23.09.2025, 5pm.

Technical specifications mentioned in tender: -	Technical specification to be read as :-
Section 4 A- Technical Specification for IVC System subhead 3 - IVC cage set, point 6 Minimum floor area of the cage should be 500 cm ² . Overall dimensions of the cages not less than 380 mm x 200mm x 140 mm (LxWxH)	Section 4 A- Technical Specification for IVC System subhead 3 - IVC cage set, point 6 Minimum floor area of the cage should be 360 cm ²
Section 4A- Technical Specification for IVC System subhead 3- IVC cage set, point 14 Water bottle should be of self-centring type and should hold a minimum of 250 ml of drinking water.	Section 4 A- Technical Specification for IVC System subhead 3- IVC cage set, point 14 Water bottle should be of self-centring type and capacity should be 200 ml or above.
Section 6- Commercial Bid, S. No. 4 - warranty 3 years (1 year + 2 Years extended warranty)	Section 6- Commercial Bid, S. No. 4 - warranty 5 years (1 year + minimum 4 Years extended warranty)

Thank you,

Prof. Sathees C. Raghavan
Chair, Central Animal Facility
IISc.