

**Supply and Installation of Solar Material Testing Facility: UV-Visible-NIR
Optical Measurement**

NOTICE INVITING GLOBAL TENDER

Department of Materials Engineering
Indian Institute of Science
Bangalore-560012
India

Date: 09/01/2026

Indian Institute of Science, Bangalore invites the best quotations from the bonafide, resourceful, and eligible manufacturer/exclusive distributor/vendors for the procurement of a Solar Material Testing Facility: UV-Visible-NIR Optical Measurement with the following technical specifications FOR IISc, Bangalore basis. The quotation should clearly mention the validity of the quote (minimum 90 days validity), terms of delivery, delivery schedule, estimated delivery date, and payment terms. The tender should be submitted in two separate sealed and distinctly marked envelopes: one containing the technical bid and the other containing the commercial bid, both of which should reach us duly signed on or before 5:00 PM, 30th January, 2026.

Table of Content

1	Section 1	Bid Schedule	
2	Section 2	Eligibility Criterion	As specified by IISc
3	Section 3	Terms and Condition	As specified by IISc
4	Section 4	Specification	Technical Specifications
5	Technical Bid	Annexure 1: Bidder Detail	
		Annexure 2: Declaration regarding experience of bidder	
		Annexure 3: Declaration regarding clean track record of the bidder.	
		Annexure 4: Declaration of acceptance of tender	
		Annexure 5: Terms and conditions. Details of item quoted	
6	Commercial bid	Quotation with price, Technical Specification of the Equipment	

1. Bid Schedule

1.	Tender Number	MT/SRR/2025/OS-1
2.	Tender Date	09/01/2026
3.	Item Description	Supply and Installation of Solar Material Testing Facility: UV-Visible-NIR Optical Measurement
4.	Tender type	Two bid system i Technical Bid (Part A) ii Commercial Bid (Part B)
5.	Place of tender submission	Office of the Chair Department of Materials Engineering Indian Institute of Science Bangalore-560012 India
6.	Last Date & Time for submission of tender	30/01/2026 by 5:00 PM
7.	For further clarification	Dr. Sachin R. Rondiya Assistant Professor Department of Materials Engineering Indian Institute of Science Bangalore-560012 India

2. Eligibility Criteria

1. This is a global tender inquiry, and any bidder with registered head offices within or outside India is eligible to bid. However, The Bidder's firm should have been in existence for a minimum of 5 years. (Enclose Company Registration Certificate).
2. Only the Original Equipment Manufacturer or their authorized representatives shall participate in the bid.
3. The order will be placed only on the bidder who participated in the bid.
4. A complete bid with the following must be submitted:
 - a) Technical bid with a technical compliance sheet, supporting documentation and **masked** commercial bid (a copy of the commercial bid, but '**without**' any pricing information) in one sealed envelope. The envelope should be '**marked**' as '**Technical Bid**'.
 - b) Commercial bid, including CIP pricing, warranty information, and other commercial terms and conditions, in another sealed envelope. The envelope should be '**marked**' as '**Commercial Bid**'.
 - c) The envelopes with technical and commercial bids should be placed in a single sealed envelope.
 - d) All envelopes must be addressed to the tenderee as per information in Point 6 of the Bid Schedule. The tender number and date must be inscribed on all the envelopes.
5. The bidder should sign and submit the declaration of Acceptance of Terms and Conditions as per **-Annexure 4**.
6. The bidder must not be blacklisted/banned/suspended or have a record of any service-related dispute with any organization in India or elsewhere. A declaration to this effect must be given as per **Annexure 3**.
7. Any bids that are deemed incomplete can be summarily rejected without further consideration.

3. Terms and Conditions

A) Submission of Tender:

1. All documentation in the tender should be in English.
2. The tender should be submitted in two envelopes (two bid system).
 - a) Technical Bid (Part-A) – Technical bid consisting of all technical details and checklist for conformance to technical specifications. The technical bid should contain a technical compliance table given in Section 4 of this document.
 - b) Commercial Bid (Part-B) – Indicating item-wise price for the items mentioned in the technical bid, as per the format of quotation provided in Annexure 6, and other commercial terms and conditions.

3. The technical and price bids should be placed in separate sealed envelopes, superscripting the tender number and the due date on both. Both of these sealed covers are to be placed in a larger envelope, which should also be sealed and duly superscripted with the Tender No., Tender Description & Due Date.
4. The SEALED COVER superscripting tender number, tender description, and due date, and the contact address, contact phone, and email of the bidder should reach **The Office, Department of Materials Engineering, Indian Institute of Science, Bangalore – 560 012, India** on or before the due date mentioned in the tender notice. Write “Supply and Installation of Solar Material Testing Facility: UV-Visible-NIR Optical Measurement” clearly on top of the envelope. If the due date falls on a holiday, the tender will be accepted on the next business day. If the quotation cover is not sealed, it will be rejected.
5. All queries are to be addressed to the person identified in “Section 1 – Bid Schedule” of the tender notice.
6. GST/other taxes, levies etc., are to be indicated separately. The bidder must mention their GST Registration and PAN in the tender document, if applicable.
7. The Institute reserves the right to accept or reject any bid, to annul the bidding process, and reject all bids at any time before the award of the contract without thereby incurring any liability to the affected bidder or bidders or any obligation to inform the affected bidder or bidders.
8. Incomplete bids will be summarily rejected.

B) Cancellation of Tender:

Notwithstanding anything specified in this tender document, IISc Bangalore, in its sole discretion, unconditionally and without having to assign any reason, reserves the rights to:

- a) Accept OR reject the lowest tender, any other tender, or all the tenders.
- b) Accept any tender in full or in part.
- c) To reject the tender, offer not confirming to the tender terms.

C) Validity of the Offer:

The offer must be valid for at least 90 Days from the closing date of this tender.

D) Evaluation of Offer:

1. The technical bid (Part A) will be opened first and evaluated. Bidders meeting the required eligibility criteria as stated in Section 2 of this document shall only be considered for Commercial Bid (Part B) opening. Further, agencies not furnishing the documentary evidence as required will not be considered.
2. Pre-qualification of the bidders shall not imply final acceptance of the Commercial Bid. The agency may be rejected at any point during technical evaluation or during commercial evaluation. The decision regarding acceptance and/or rejection of any offer in part or full shall be at the sole discretion of IISc Bangalore, and the decision in this regard shall be binding on the bidders.
3. The award of the contract will be subject to acceptance of the terms and conditions stated in this tender.
4. Any offer which deviates from the vital conditions (as illustrated below) of the tender is liable to be rejected:

- a) Non-submission of complete offers.
- b) Receipt of bids after the due date and time and/ or by email/fax (unless specified otherwise).
- c) Receipt of bids in open conditions.

5. In case any BIDDER is silent on any clauses mentioned in these tender documents, IISc Bangalore shall consider that the BIDDER had accepted the clauses as of the tender and no further claim will be entertained. Further if the BIDDER is silent or does not give detail justification of their claim regarding those mentioned in technical specifications, IISc Bangalore reserves the full right to reject the tender due to non-compliance without any further discussion.

6. No revision in the terms and conditions quoted in the offer will be entertained after the last date and time fixed for receipt of tenders.

7. Lowest bid will be calculated based on the total price of all items tendered for Basic equipment along with accessories selected for installation, operation, preprocessing and post processing, optional items, recommended spares, warranty and annual maintenance contract. The purchase committee seeks the most cost -effective solution for obtaining a new tool. Vendors are encouraged to propose all avenues, including but not limited to buy back of the existing tool, turnkey upgrade of existing to, 1 or purchase of a new tool.

E) Pre-requisites for Installation:

The bidder must provide the prerequisite installation requirements of the equipment along with the technical bid.

F) Warranty:

The complete system is to be under warranty **period of 1 year (comprehensive)** from the date of functional installation. The vendor should include the cost of any spares needed during the warranty period, including electronics, sub-components, and software. If the instrument is found to be defective, it has to be replaced or rectified at the cost of the bidder within 30 days from the date of receipt of written communications from IISc, Bangalore.

1. If there is any delay in replacement or rectification, the warranty period should be correspondingly extended.
2. Terms and conditions for the annual maintenance contract beyond the warranty period should be mentioned.
3. Warranty terms and additional warranty options are a must for all the components. Specify the service plan, like whether the local distributor will address the issue or the parent company. A minimum of three years of complete system warranty (comprehensive and non-comprehensive) should be given. If the system requires service during the warranty period, the vendor must guarantee or replace the instrument for free. Vendor to have logistic support to ensure that at least 95 % of the service parts are readily available and upkeep delivery within 1 week.
4. A declaration of Conformity certificate and document that provides information on testing and validation must be provided. All modules must be GLP compliant.
5. Support should be available on full working days (excluding Public Holidays), local time.
6. On-site installation, commissioning, and training shall be conducted by a qualified factory trained engineer.
7. The vendor must demonstrate that it has a proven appropriate set-up and capability to provide after-sales service efficiently and effectively. The supplier should have a similar system in their

facility to that proposed in this tender for training purposes.

8. The vendor must have a local dedicated Sales & Service team & Application lab in the Southern region.

F.1) Annual Maintenance Contract:

An annual maintenance contract for at least two years post-warranty may be provided as an essential, optional item upon completion of the warranty period.

F.2) SPARES:

Vendors must provide a detailed list of spares and a user manual with a detailed Bill of Materials for all Parts. It should include the Spares Column with the Manufacturer part Number, Qty, and availability of stock after 3 Years.

G) Purchase Order:

1. The order will be placed on the bidder whose bid is accepted by IISc based on the terms & conditions mentioned in the tender document.
2. The quantity of the items in the tender is only indicative. IISc, Bangalore, reserves the right to increase /decrease the quantity of the items depending on the requirement.
3. If the quality of the product and service provided is not satisfactory, IISc Bangalore reserves the right to cancel or amend the contract.

H) Delivery, Installation, and Training:

The bidder shall provide the lead time to delivery, installation and made functional at IISc, Bangalore from the date of issuance of purchase order (PO). The system should be delivered, installed and made functional within 120 days from the date of receipt of PO. The supply of the items will be considered as effected only on satisfactory installation and inspection of the system and inspection of all the items and features/capabilities tested by the IISc, Bangalore. After successful installation and inspection, the date of taking over of entire system by the IISc, Bangalore shall be taken as the start of the warranty period. No partial shipment is allowed. The bidder should also arrange for technical training to the local facility technologists and users.

I) Payment Terms:

No advanced payment will be made for Indigenous purchase. However, 90% Payment against Delivery and 10 % after installation are agreed to wherever the installation is involved. Any loss due to fluctuation in foreign exchange rates will be at the beneficiary account. AMC cost (if ordered), after completion of warranty period) will be released on half-yearly basis at the end of each six months subject to satisfactory services. Price basis must be on IISc, Bangalore basis only. As per GFR, no advance payment can be made to domestic vendors unless an equal amount of bank guarantee is provided.

J) Statutory Variation:

Any statutory increase in the taxes and duties subsequent to bidder's offer, if it takes place within the original contractual delivery date, will be borne by IISc, Bangalore, subject to the claim being supported by documentary evidence. However, if any decrease takes place, the advantage will have to be passed onto IISc, Bangalore.

K) Disputes and Jurisdiction:

Any legal disputes pertaining to this tender or any breach of contract shall be settled in the court of competent jurisdiction located within the city of Bangalore, India.

L) General:

1. All amendments, time extension, clarifications etc., within the period of submission of the tender will be communicated electronically. No extension in the bid due date/time shall be considered on account of delay in receipt of any document(s) by mail.
2. The bidder may furnish any additional information, which is necessary to establish capabilities to successfully complete the envisaged work. It is, however, advised not to furnish superfluous information. With prior intimation, the bidder may visit the installation site before tender submission.
3. Any information furnished by the bidder found to be incorrect, either immediately or at a later date, would render the bidder liable to be debarred from tendering/taking up of work in IISc, Bangalore.
4. Operation and service manual in English (electronic and hard copy) should be provided. A document that explains the list of items and their placement in the system (e.g., a schematic diagram) for future reference. Please note that this is not related to your intellectual property (IP). We require maximum information about the components, such as detectors and connections, which will be helpful in the future.
5. Standard samples (if required) to be provided by the company for testing the instruments at the time of installation on-site to the quoted accuracy in the given technical specification for the demonstration of the performance of equipment.
6. The vendor should have a track record of having previously supplied at least five identical instruments in CFTIs such as JNCASR, IITs, IISERs, and NITs with the specifications mentioned in the current tender. Details of such systems should be provided. The vendor must provide the end-user list (with contact details including emails and phone numbers) of at least 5 customers from Indian Institutes /Labs with contact person name, address, phone, fax, and email IDs, certificates confirming satisfactory performance. The primary focus of these installed systems should have included reliable data in the form of pictorial graphs must be provided.
8. The committee reserves the right to reject the technical bid if the above condition is not satisfied
9. Wherever requested data must be supplied along with technical compliance documents, technical bids without supporting data will be deemed technically non-compliant.
10. Price of every line item in the commercial bid should be quoted along with the total quoted price for the instrument to be operational (installed and ready to use) in our facility. Quote the price of each optional line item separately.
11. The vendor should have qualified technical service personnel for the equipment based in India (preferably in Bangalore).
12. Items in addition to that listed in the technical table that you would like to bring to our attention, such as data sheets, technical plots, etc. can be listed at the end of the compliance table
13. Vendors are encouraged to highlight the advantages of their instrument and accessories over comparable instruments from competitors.
14. If needed, a meeting for any technical clarifications can be scheduled with the undersigned by sending an email.
15. The Institute reserves the right to accept or reject any bid or to annul the bidding process and reject all bids at any time prior to the award of the contract without thereby incurring any liability of the affected bidder or bidders.

16. After the award of the purchase order, the vendor must provide an Order Acknowledgment within 7 days from the receipt of the Purchase Order.
17. The vendor must have a local dedicated Sales & Service team & Application lab in the Southern region.
18. Vendors must provide proper justification for any technical deviations mentioned in the technical comparison statement during evaluation.
19. A comprehensive one-year warranty must cover all system components and accessories supplied with the equipment.
20. Vendors must submit a detailed list of infrastructure requirements (such as power supply, exhaust, laboratory space, table dimension etc.) necessary for installation and smooth operation of the system.
21. The payment terms should be specified in the commercial proposal, which should be consistent with IISc's purchase policies.
22. Provide details of the number of trained personnel in India, the number in the southern region, or Bangalore who can service the instrument.
23. Include other options currently available which can be added in the future.
24. The vendor should attach product brochures along with the technical bid.

4. Technical Specification

<u>Technical Specifications for Solar Material Testing Facility: UV-Visible-NIR Optical Measurement</u>	
Configuration Design: Solar Material Testing Facility: UV-Visible-NIR Optical Measurement	
General Description	
Essential	<ol style="list-style-type: none"> 1. PC controlled Solar Material Testing Facility: UV-Visible-NIR Optical Measurement should work in transmission, Reflection and Absorption mode with high accuracy and precision. Czerny -Turner mount, Double beam, high performance blazed holographic grating-grating type double monochromator system to ensure high resolution and low stray light level across the wide measurement range. 2. Instrument must incorporate factory pre-aligned Deuterium and Tungsten-Halogen sources. Automatic adjustment should be present inside the instrument for automated wavelength accuracy validation and light source position. Automatic switching synchronized with wavelength switching selectable between 282 – 393 nm range (0.1 nm increments or better). 3. The Solar Material Testing Facility: UV-Visible-NIR Optical Measurement must be fitted with three detectors to achieve high sensitivity in the NIR range. 4. NIR quartz cell, powder holder should be supplied with instrument. 5. Software operation should be user friendly with controlled operation, data collection and analysis. 6. The system should be plug and play type and easy to install or to be removed. 7. The system has to be optimized for standard/reference sample and to be successfully demonstrated at our site. 8. The instrument must allow standard RS232C or USB communication with PC. 9. All cables and connectors to be provided for full operation of the system 10. Complete product catalog describing all the required basic and optional items should be produced. 11. Functions should be available to introduce temperature dependent upgrade accessories in future. 12. OEM should have application lab and service facility in India for support. 13. The instrument should be flexible to incorporate spare-parts to enhance operation range and future expandability to support a wide range of experimental needs.
Technical Specification	
Essential	<ol style="list-style-type: none"> 1. Wavelength Range: 185 nm - 3300 nm (Minimum or better). 2. Detector: PMT, InGaAs and Cooled PbS 3. Wavelength accuracy: ± 0.2 nm in UV-Vis range and ± 0.8 nm in NIR region (Minimum or better).

	<p>4. Wavelength reproducibility: ± 0.08 nm in UV-Vis range and ± 0.3 nm in NIR region (Minimum or better).</p> <p>5. Spectral Bandwidth: 8 steps (0.1 nm - 8 nm) in UV-Vis region and 10 steps (0.2 nm - 32 nm) in NIR range (Minimum or better).</p> <p>6. Resolution: ± 0.1 nm in UV-Vis range and ± 0.5 nm in NIR region (Minimum or better).</p> <p>7. Stray light (T%): 0.00008% (220 nm, NaI), 0.00005% (340 nm, NaNO₂), 0.0005% (1420 nm, H₂O), 0.005% (2365 nm, CHCl₃) (Minimum or better).</p> <p>8. Baseline flatness: ± 0.001 A (Minimum or better)</p> <p>9. Wavelength Sampling Pitch: 0.01 nm (Minimum or better)</p> <p>10. Baseline stability: 0.0002 Abs/hr at 500 nm (Minimum or better).</p> <p>11. Scan speed: 4500 nm/min (UV-Vis), 4000 nm/min (NIR PbS), 9000 nm/min (NIR-PMT/InGaAs) (Minimum or better).</p> <p>12. Photometric range: -6 to +6 Abs in UV-Vis-NIR range (Minimum or better) in double beam photometric system.</p> <p>13. Photometric accuracy (using NIST filter): ± 0.003 (1A), ± 0.002 (0.5A) (Minimum or better).</p> <p>14. Photometric Noise: 0.00005 Abs (500 nm), 0.00008 Abs (900 nm), 0.00003 Abs (1500 nm) (Minimum or better)</p>
	<p>Integrating Sphere specification:</p> <ol style="list-style-type: none"> 1. Size: Sample chamber should accommodate with 60 mm inner diameter Integrating Sphere (minimum or better) for best suitable result. All type of sample holders (Solid, Liquid and Film) for Integrating sphere should be provided for measurement of diffuse/specular reflectance of solid samples and transmission of liquid or solid samples. 2. Detector: Must be Combination of three detectors – PMT, InGaAs and PbS. 3. Wavelength range: 220 nm – 2600 nm (Minimum or better) 4. Standard material BaSO₄ for integrating sphere should be provided.
software specifications	
	<ul style="list-style-type: none"> ▪ Software Operation should be user friendly ▪ Measurement modes: Spectrum, quantitation, photometric, and time course Save data files, parameter files, and template files, Retain history of changes to data files and parameter files. Manage sample information (sample name, sample ID, comments, etc.). Specify all sample information settings before measurements. Control automatically from external application. Real-time display of wavelength, photometric value, and concentration values Graph settings (line type, line colour, etc.), Adjust graph scale or use auto-scale. Automatically send measurement data to Excel spreadsheet. • Automatically output measurement data in text format. automatically analyze data after measurements (evaluation function, peak detection, extract photometric value for specified wavelength, calculate area, correction, and conversion). Automatically print report after measurements. Overlay spectral waveforms. Analysis and pass/fail judgment using spectral evaluation function

	<p>Data processing (detect peaks, extract photometric value of specified wavelength, calculate area) Correction (dilution factor correction, optical path length correction, etc.) Conversion (smoothing, differentiation, etc.)</p> <ul style="list-style-type: none"> ▪ Quantitation for specified wavelengths (one wavelength, difference between two wavelengths, ratio of two wavelengths, three wavelengths) Quantitation based on maximum/area spectrum value Single-point calibration curve, multi-point calibration curve, K-factor method Calibration curve method (first to third-order equations) Correct dilution factor for each sample. Specify weighting factors for each sample (standard samples). Concentration value pass/fail judgment Perform repeated measurements
--	--

<u>Installation & Commissioning</u>	
Essential	<ul style="list-style-type: none"> ➢ Installation and commissioning of the equipment has to be carried out by supplier at our laboratory in site and the performance has to be demonstrated. ➢ Onsite training must be provided to our personnel on the installed equipment for operation and data processing at free of cost.
<u>Instrument Control</u>	
Essential	<ul style="list-style-type: none"> ➢ A Desktop computer, Windows Operating System (OEM) shall be provided with required Configuration.
<u>Warranty</u>	
Essential	System shall be warranted for a minimum period of 1 year from the date of installation of the equipment at site.
<u>Maintenance</u>	
Essential	<ul style="list-style-type: none"> ➢ The system Should be plug & play type and easy to install or to be removed. ➢ OEM should have application lab facility in India for support. ➢ Service facilities should be available in Karnataka. ➢ Please provide service center details along with contact numbers in India.

5. Technical Bid

The technical bid should furnish all requirements of the tender along with all annexures in this section and submitted to

The Chairperson,
Attn: Dr. Sachin R. Rondiya (Assistant Professor)
Department of Materials Engineering,
Indian Institute of Science
Bangalore – 560012, India

Annexure - 1

Details of the Bidder

The bidder must provide the following mandatory information & attach supporting documents wherever mentioned:

Sl. No.	Items	Details
1	Name of the Bidder	
2	Nature of Bidder (Attach attested copy of Certificate of Incorporation/ Partnership Deed)	
3	Registration No/ Trade License, (attach attested copy)	
4	Registered Office Address	
5	Address for communication	
6	Contact person- Name and Designation	
7	Telephone No	
8	Email ID	
9	Website	
10	PAN No. (attach copy)	
11	GST No. (attach copy)	

Signature of Bidder

Name

Designation, Seal

Date:

Annexure - 2

Declaration regarding experience

To,
The Chairperson,
Materials Engineering Department,
Indian Institute of Science
Bangalore – 560012, India

Ref: Tender No: MT/SRR/2025/OS-1

Dated: 09/01/2026

Supply and Installation of Solar Material Testing Facility: UV-Visible-NIR Optical Measurement

Dear Sir,

I've carefully gone through the Terms & Conditions contained in the above referred tender. I hereby declare that my company / firm has _____ years of experience in Supply and Installation of Solar Material Testing Facility: UV-Visible-NIR Optical Measurement with asked technical specifications.

(Signature of the Bidder)

Printed Name

Designation, Seal

Date:

Annexure-3

Declaration regarding track record

To,
The Chairperson,
Materials Engineering Department,
Indian Institute of Science
Bangalore – 560012, India

Ref: Tender No: MT/SRR/2025/OS-1

Dated: 09/01/2026

Supply and Installation of Solar Material Testing Facility: UV-Visible-NIR Optical Measurement

Dear Sir,

I've carefully gone through the Terms & Conditions contained in the above referred tender. I hereby declare that my company/ firm is not currently debarred /blacklisted by any Government / Semi Government organizations / institutions in India or abroad. I further certify that I'm competent officer in my company / firm to make this declaration.

Or I declare the following

Sl. No.	Country in which the company is Debarred /blacklisted / case is Pending	Blacklisted / debarred by Government / Semi Government/Organizations /Institutions	Reason	Since when and for how long

(NOTE: In case the company / firm was blacklisted previously, please provide the details regarding period for which the company / firm was blacklisted and the reason/s for the same).

Yours faithfully

(Signature of the Bidder)

Name

Designation, Seal

Date:

Annexure-4

Declaration for acceptance of terms and conditions

To,
The Chairperson,
Materials Engineering Department,
Indian Institute of Science
Bangalore – 560012, India

Ref: Tender No: Tender No: MT/SRR/2025/OS-1

Dated: 09/01/2026

Supply and Installation of Solar Material Testing Facility: UV-Visible-NIR Optical Measurement

Dear Sir,

I've carefully gone through the Terms & Conditions as mentioned in the above referred tender document. I declare that all the provisions of this tender document are acceptable to my company. I further certify that I'm an authorized signatory of my company and am, therefore, competent to make this declaration.

Yours faithfully,

(Signature of the Bidder)

Name

Designation, Seal

Date:

Annexure-5

Details of items quoted:

a. Company Name	
b. Product Name	
c. Part/Catalogue Number	
d. Product description / main features	
e. Detailed technical specifications	
f. Remarks	

Instructions to bidders:

1. Bidder should provide technical specifications of the quoted product/s in detail.
2. Bidder should attach product brochures along with technical bid.
3. Bidders should clearly indicate compliance or non-compliance of the technical specifications provided in the tender document.

6. Commercial bid

The commercial bid should be furnished with all requirements of the tender with supporting documents as mentioned under:

Sl. No.	Description	Cat. Number	Quantity	Unit price	Sub total
1.	Essential items noted in the technical specification				
1.a	... (details of essential items)				
1.b	...				
2.	Optional items noted in the technical specification				
2.a(details of optional items)				
2.b				
3.	Accessories for operation and installation				
4.	All Consumables, spares and software to be supplied locally				
5.	Warranty (1 year)				
6.	AMC 2 years beyond warranty				

Any additional item

Sl. No.	Description	Cat. Number	Quantity	Unit Price	Sub total

Addressed to
The Chairperson,
Attn: Dr. Sachin R. Rondiya (Assistant Professor)
Materials Engineering, Indian Institute of Science
Bangalore – 560012, India

7. Checklist

(This should be enclosed with technical bid- Part A)

The following items must be checked before the Bid is submitted:

1. Sealed Envelope “A”: Technical Bid

1. Section 5- Technical Bid (each page signed by the authorized signatory and sealed) with the below annexures:

Annexure 1: Bidders details

Annexure 2: Declaration regarding experience

Annexure 3: Declaration regarding clean track record

Annexure 4: Declaration for acceptance of terms and conditions

Annexure 5: Details of items quoted

2. Copy of this tender document duly signed by the authorized signatory on every page and sealed.

2. Sealed Envelope “B”: Commercial Bid

Section 6: Commercial Bid

Your quotation must be submitted in two envelopes: Technical Bid (Envelope A) and Commercial Bid (Envelope B) super scribbling on both the envelopes with Tender No. and due date and both of these in sealed covers and put in a bigger cover which should also be sealed and duly super scribbled with Tender No., Tender description & Due Date.