



CORRIGENDUM

Tender Reference: IISc-Med-2025-26/L-40

Date: 09/03/2026

Tender Name : Request for Quotation (RFQ) – Turnkey Central Sterile Services Department (CSSD)

Local Tender: Turnkey Central Sterile Services Department (CSSD)

All prospective bidders are hereby informed of the following clarification regarding the above-mentioned tender:

Bidders may quote the Endoscope Washer Disinfector with multi-scope capacity (2 / 3 scopes) as an optional item, and not as a standard configuration.

Technical Specification – Endoscope Washer Disinfector (Scope of supply reference-6)

1. Capacity: The system shall have the capacity to process at least 2 / 3 endoscopes simultaneously

2. Volume: The bidder shall specify the chamber volume in the technical bid.

All other terms and conditions of the tender remain unchanged.

The deadline for submission of proposals remains 11 March 2026 (Wednesday) at 05:30 PM IST.

IISc Bangalore