



CORRIGENDUM

Tender Reference number: IISc-Med-2026-27/G-3 dated 23rd April 2026 .

Date: 30/04/2026

Global tender: Planning, supply, installation, testing, commissioning, and user training of an Ultra-Low Height Radiolucent Surgical Table with Robotic Compatibility

Tender specification		Corrigendum to tender specification	
SN	Specification	SN.	Specification
A	Ultra-low height radiolucent surgical table with robotic compatibility	A	Ultra-low height radiolucent surgical table with robotic compatibility
3	The operating table shall be electrohydraulic, enabling all table movements.	3	The operating table shall be electrohydraulic/electromechanical, enabling all table movements.

All other specifications, terms and conditions remain unchanged.

IISc Bangalore