



## **CORRIGENDUM**

**Tender number: IISc-Med-2026-27/G-5 Dated -30/04/26**

Date: 06/05/2026

Global tender: Supply, installation, testing, commissioning, and comprehensive user training of a complete Flexible Video Bronchoscope.

**TENDER SPECIFICATION CORRIGENDUM TO EARLIER REQUIREMENT-(pg number-15)**

<b>Tender specification</b>		<b>Corrigendum to tender specification</b>	
<b>SN</b>	<b>Specification</b>	<b>SN.</b>	<b>Specification</b>
<b>B</b>	<b>Flexible Video Bronchoscope</b>	<b>B</b>	<b>Flexible Video Bronchoscope</b>
4	The scope insertion tube outer diameter should be 6mm or more	4	The scope insertion tube outer diameter should be equal to 5mm or more.

All other specifications, terms and conditions remain unchanged.

IISc Bangalore